

LB130681


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48800



A= 0.0000e+000

B= 2.0390e-004

C= -7.0348e-003

Rho= 0.9997786

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Good
0.0	0.000	0.076	0.076	406	0.000	406					
0.2	0.200	0.183	-0.017	934	0.0 %	934					
2.5	2.500	2.367	-0.133	11643	0.0 %	11643					
5.0	5.000	5.045	0.045	24777	0.0 %	24777					
7.5	7.500	7.585	0.085	37234	0.0 %	37234					
10.0	10.000	9.944	-0.056	48800	0.0 %	48800					

Good
 1
 1
 1
 1
 1
 1

LB130681 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 May 2024 10:59:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	406	-	0.0000	7471B	PPB	09 May 2024 11:06:33	S	Std
0.2 - 1	50.2	934	-	0.2000	7471B	PPB	09 May 2024 11:08:50	S	Std
2.5 - 1	5.25	11643	-	2.5000	7471B	PPB	09 May 2024 11:11:07	S	Std
5.0 - 1	55	24777	-	5.0000	7471B	PPB	09 May 2024 11:13:25	S	Std
7.5 - 1	57.5	37234	-	7.5000	7471B	PPB	09 May 2024 11:15:44	S	Std
10.0 - 1	58.6	48800	-	10.0000	7471B	PPB	09 May 2024 11:20:51	S	Std
ICV39 - 1	ICV39	20335	4.1394	-	7471B	PPB	09 May 2024 11:23:33	U	SMPL
ICB39 - 1	ICB39	119	0.0172	-	7471B	PPB	09 May 2024 11:25:50	U	SMPL
CCV55 - 1	CCV55	26440	5.3842	-	7471B	PPB	09 May 2024 11:28:11	U	SMPL
CCB55 - 1	CCB55	133	0.0201	-	7471B	PPB	09 May 2024 11:30:26	U	SMPL
CRA - 1	CRA	894	0.1753	-	7471B	PPB	09 May 2024 11:32:48	U	SMPL
HighStd - 1	HighStd	49855	10.1586	-	7471B	PPB	09 May 2024 11:35:03	U	SMPL
ChkStd - 1	ChkStd	36320	7.3988	-	7471B	PPB	09 May 2024 11:37:18	U	SMPL
PB160795BL - 1	PBS	208	0.0354	-	7471B	PPB	09 May 2024 11:42:26	U	SMPL
PB160795BS - 1	LCSS	18106	3.6849	-	7471B	PPB	09 May 2024 11:48:21	U	SMPL
P2312-01 - 1	B-125-SB-01	14385	2.9261	-	7471B	PPB	09 May 2024 11:50:37	U	SMPL
P2312-02 - 1	B-125-SB-02	5921	1.2003	-	7471B	PPB	09 May 2024 11:52:57	U	SMPL
P2375-01 - 1	WC-A-COMP	8625	1.7516	-	7471B	PPB	09 May 2024 11:55:17	U	SMPL
P2395-01 - 1	SP-DRIVEWAY	951	0.1869	-	7471B	PPB	09 May 2024 11:57:36	U	SMPL
CCV56 - 1	CCV56	24893	5.0688	-	7471B	PPB	09 May 2024 12:02:09	U	SMPL
CCB56 - 1	CCB56	285	0.0511	-	7471B	PPB	09 May 2024 12:04:24	U	SMPL
P2395-01MS - 1	SP-DRIVEWAYMS	21115	4.2984	-	7471B	PPB	09 May 2024 12:06:44	U	SMPL
P2395-01MSD - 1	SP-DRIVEWAYMSD	23713	4.8282	-	7471B	PPB	09 May 2024 12:08:59	U	SMPL
P2395-04 - 1	SP-INSIDE-GATE	1212	0.2401	-	7471B	PPB	09 May 2024 12:11:18	U	SMPL
P2398-01 - 1	WC-KK	837	0.1636	-	7471B	PPB	09 May 2024 12:13:38	U	SMPL
P2404-02 - 1	TOP-SOIL-1	3818	0.7715	-	7471B	PPB	09 May 2024 12:15:54	U	SMPL
P2404-04 - 1	TOP-SOIL-2	19758	4.0217	-	7471B	PPB	09 May 2024 12:18:10	U	SMPL
P2404-06 - 1	TOP-SOIL-3	3848	0.7776	-	7471B	PPB	09 May 2024 12:20:28	U	SMPL
P2408-01 - 1	BIN-1	3778	0.7633	-	7471B	PPB	09 May 2024 12:22:47	U	SMPL
P2408-03 - 1	ETGI-285	2265	0.4548	-	7471B	PPB	09 May 2024 12:25:06	U	SMPL
P2411-01 - 1	WCS-TPA	18923	3.8515	-	7471B	PPB	09 May 2024 12:27:23	U	SMPL
CCV57 - 1	CCV57	26976	5.4935	-	7471B	PPB	09 May 2024 12:29:40	U	SMPL
CCB57 - 1	CCB57	299	0.0539	-	7471B	PPB	09 May 2024 12:31:58	U	SMPL
P2412-01 - 1	SS-1-COMP	6179	1.2529	-	7471B	PPB	09 May 2024 12:34:18	U	SMPL
P2413-01 - 1	WC-CC	2542	0.5113	-	7471B	PPB	09 May 2024 12:36:33	U	SMPL
P2413-05 - 1	WC-EE	1637	0.3268	-	7471B	PPB	09 May 2024 12:38:51	U	SMPL
P2414-01 - 1	1011UMR-SS004-0212-01	10751	2.1851	-	7471B	PPB	09 May 2024 12:41:08	U	SMPL
P2415-01 - 1	1011UMR-SS005-0212-01	5519	1.1183	-	7471B	PPB	09 May 2024 12:43:23	U	SMPL
P2395-01DUP - 1	SP-DRIVEWAYDUP	1093	0.2158	-	7471B	PPB	09 May 2024 12:49:18	U	SMPL
P2395-01LX5 - 1		188	0.0313	-	7471B	PPB	09 May 2024 12:51:34	U	SMPL
P2395-01A - 1		20670	4.2077	-	7471B	PPB	09 May 2024 12:53:51	U	SMPL
CCV58 - 1	CCV58	25164	5.1240	-	7471B	PPB	09 May 2024 12:56:08	U	SMPL
CCB58 - 1	CCB58	216	0.0370	-	7471B	PPB	09 May 2024 12:58:26	U	SMPL