

CB130889


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

4498E



A= 0.0000e+000

B= 2.1931e-004

C= 4.6251e-002

Rho= 0.9992552

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.074	0.074	128	0.000	128					400
0.2	0.200	0.223	0.023	806	0.0 %	806					12
2.5	2.500	2.435	-0.065	10890	0.0 %	10890					-3
5.0	5.000	4.802	-0.198	21685	0.0 %	21685					-4
7.5	7.500	7.754	0.254	35145	0.0 %	35145					3
10.0	10.000	9.912	-0.088	44988	0.0 %	44988					-1

LB130689 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 May 2024 14:26:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	128	-	0.0000	7470A	PPB	09 May 2024 14:31:27	S	Std
0.2 - 1	50.2	806	-	0.2000	7470A	PPB	09 May 2024 14:33:44	S	Std
2.5 - 1	52.5	10890	-	2.5000	7470A	PPB	09 May 2024 14:36:01	S	Std
5.0 - 1	85	21685	-	5.0000	7470A	PPB	09 May 2024 14:38:18	S	Std
7.5 - 1	87.5	35145	-	7.5000	7470A	PPB	09 May 2024 14:40:35	S	Std
10.0 - 1	510	44988	-	10.0000	7470A	PPB	09 May 2024 14:48:06	S	Std
ICV40 - 1	ICV40	17972	3.9876	-	7470A	PPB	09 May 2024 14:51:02	U	SMPL
ICB40 - 1	ICB40	-248	-0.0081	-	7470A	PPB	09 May 2024 14:53:18	U	SMPL
CCV59 - 1	CCV59	22677	5.0195	-	7470A	PPB	09 May 2024 14:55:36	U	SMPL
CCB59 - 1	CCB59	-360	-0.0327	-	7470A	PPB	09 May 2024 14:57:51	U	SMPL
CRA - 1	CRA	614	0.1809	-	7470A	PPB	09 May 2024 15:00:08	U	SMPL
HighStd - 1	HighStd	44460	9.7966	-	7470A	PPB	09 May 2024 15:02:23	U	SMPL
ChkStd - 1	ChkStd	31760	7.0114	-	7470A	PPB	09 May 2024 15:04:38	U	SMPL
PB160828BL - 1	PBW	-574	-0.0796	-	7470A	PPB	09 May 2024 15:06:58	U	SMPL
PB160828BS - 1	LCSW	19626	4.3504	-	7470A	PPB	09 May 2024 15:09:16	U	SMPL
P2335-02 - 1	COMP	806	0.2230	-	7470A	PPB	09 May 2024 15:11:33	U	SMPL
P2410-01 - 1	A2295	327	0.1180	-	7470A	PPB	09 May 2024 15:13:52	U	SMPL
P2417-02 - 1	COMP	516	0.1594	-	7470A	PPB	09 May 2024 15:16:11	U	SMPL
P2417-02DUP - 1	COMPDUP	690	0.1976	-	7470A	PPB	09 May 2024 15:18:29	U	SMPL
P2417-02MS - 1	COMPMS	11839	2.6426	-	7470A	PPB	09 May 2024 15:20:46	U	SMPL
CCV60 - 1	CCV60	22303	4.9375	-	7470A	PPB	09 May 2024 15:25:42	U	SMPL
CCB60 - 1	CCB60	-398	-0.0410	-	7470A	PPB	09 May 2024 15:27:58	U	SMPL
P2417-02MSD - 1	COMPMSD	11234	2.5099	-	7470A	PPB	09 May 2024 15:30:15	U	SMPL
P2418-01 - 1	1410	-175	0.0079	-	7470A	PPB	09 May 2024 15:32:30	U	SMPL
P2417-02LX5 - 1		2	0.0467	-	7470A	PPB	09 May 2024 15:34:49	U	SMPL
P2417-02A - 1		16366	3.6354	-	7470A	PPB	09 May 2024 15:37:04	U	SMPL
CCV61 - 1	CCV61	24280	5.3710	-	7470A	PPB	09 May 2024 15:39:20	U	SMPL
CCB61 - 1	CCB61	-449	-0.0522	-	7470A	PPB	09 May 2024 15:41:37	U	SMPL