

Analytical Summary Report

Analysis Method: 9040C
 Parameter: pH
 Run Number: LB130702

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/10/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/10/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	05/10/2024	08:20
4	ICV	1	Water	NA	NA	20.4	7.01	05/10/2024	08:22
5	CCV1	1	Water	NA	NA	20.3	2.02	05/10/2024	08:25
6	P2418-01	1	Water	NA	NA	22.4	9.24	05/10/2024	08:35
7	P2418-01DUP	1	Water	NA	NA	22.5	9.25	05/10/2024	08:36
8	P2436-01	1	Water	NA	NA	23.6	5.61	05/10/2024	08:44
9	P2436-02	1	Water	NA	NA	24.0	5.38	05/10/2024	08:48
10	P2436-03	1	Water	NA	NA	24.1	5.36	05/10/2024	08:50
11	P2437-06	1	Water	NA	NA	22.2	5.52	05/10/2024	09:00
12	P2437-07	1	Water	NA	NA	21.5	5.99	05/10/2024	09:10
13	P2437-08	1	Water	NA	NA	25.2	6.99	05/10/2024	09:15
14	P2439-01	1	Water	NA	NA	20.2	5.54	05/10/2024	09:25
15	P2440-03	1	Water	NA	NA	20.6	6.23	05/10/2024	09:30
16	CCV2	1	Water	NA	NA	20.3	12.02	05/10/2024	09:35

WORKLIST(Hardcopy Internal Chain)

B0702

WorkList Name : PH P2440 **WorkList ID :** 180139 **Department :** Wet-Chemistry **Date :** 05-10-2024 08:03:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2418-01	1410	Water	pH	Cool 4 deg C	PSEG03	K51	05/08/2024	9040C
P2436-01	CELL-A	Water	pH	Cool 4 deg C	STER09	K51	05/07/2024	9040C
P2436-02	CELL-B/C	Water	pH	Cool 4 deg C	STER09	K51	05/07/2024	9040C
P2436-03	CELL-D	Water	pH	Cool 4 deg C	STER09	K51	05/07/2024	9040C
P2437-06	OUTSIDE-BLACK-DRUM	Water	pH	Cool 4 deg C	PSEG03	K41	05/09/2024	9040C
P2437-07	OUTSIDE-RED-DRUM	Water	pH	Cool 4 deg C	PSEG03	K41	05/09/2024	9040C
P2437-08	WAREHOUSE-TOTE-WATER	Water	pH	Cool 4 deg C	PSEG03	K41	05/09/2024	9040C
P2439-01	50319	Water	pH	Cool 4 deg C	PSEG03	L11	05/09/2024	9040C
P2440-03	EQ-TANK-CONTAINMENT-PIT	Water	pH	Cool 4 deg C	PSEG03	L11	05/09/2024	9040C

Date/Time 05.10.24 08:10 **Date/Time** 05.10.24 12:10
Raw Sample Received by: JWC (Sum) **Raw Sample Received by:** J.C. Sum
Raw Sample Relinquished by: JWC (Sum) **Raw Sample Relinquished by:** JWC