


L3130708

SFAM01.1

INSTRUMENT ID: (V)

Linear



A= 0.0000e+000

B= 2.1783e-004 slope

C= 3.2377e-002 Intercept

Rho= 0.9999080

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/oD
0.00	0.000	0.048	0.048	73	0.000	73					1
0.05	0.050					442	0				14
0.20	0.200	0.227	0.027	895	0.0 %	895					2
2.50	2.500	2.452	-0.048	11107	0.0 %	11107					2
5.00	5.000	4.912	-0.088	22400	0.0 %	22400					0
7.50	7.500	7.526	0.026	34403	0.0 %	34403					0
10.0	10.000	10.034	0.034	45915	0.0 %	45915					

LB130708

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	73	-		0 SFAM01.1	PPB	5/10/2024 12:08	Std
	0.2 S01	895	-		0.2 SFAM01.1	PPB	5/10/2024 12:10	Std
	2.5 S02	11107	-		2.5 SFAM01.1	PPB	5/10/2024 12:13	Std
	5 S03	22400	-		5 SFAM01.1	PPB	5/10/2024 12:15	Std
	7.5 S04	34403	-		7.5 SFAM01.1	PPB	5/10/2024 12:17	Std
	10 S05	45915	-		10 SFAM01.1	PPB	5/10/2024 12:22	Std
ICV041	ICV041	18362	4.0322 -		SFAM01.1	PPB	5/10/2024 12:25	SMPL
ICB041	ICB041	-286	-0.0299 -		SFAM01.1	PPB	5/10/2024 12:27	SMPL
CCV062	CCV062	22901	5.021 -		SFAM01.1	PPB	5/10/2024 12:30	SMPL
CCB062	CCB062	-270	-0.0264 -		SFAM01.1	PPB	5/10/2024 12:32	SMPL
PB160854BL	PBW864	-30	0.0258 -		SFAM01.1	PPB	5/10/2024 12:34	SMPL
P2205-01	E1D02	-32	0.0254 -		SFAM01.1	PPB	5/10/2024 12:36	SMPL
P2205-08	E1CZ0	-9	0.0304 -		SFAM01.1	PPB	5/10/2024 12:39	SMPL
P2205-17	E1D03	11	0.0348 -		SFAM01.1	PPB	5/10/2024 12:41	SMPL
P2205-18	E1CZ1	8	0.0341 -		SFAM01.1	PPB	5/10/2024 12:43	SMPL
P2205-19	E1CZ2	6	0.0337 -		SFAM01.1	PPB	5/10/2024 12:45	SMPL
P2205-20	E1D04	13	0.0352 -		SFAM01.1	PPB	5/10/2024 12:48	SMPL
P2267-01	YCLC5	3522	0.7996 -		SFAM01.1	PPB	5/10/2024 12:50	SMPL
P2267-02	YCLC5D	3813	0.863 -		SFAM01.1	PPB	5/10/2024 12:52	SMPL
P2267-03	YCLC5S	8427	1.8681 -		SFAM01.1	PPB	5/10/2024 12:55	SMPL
P2268-01	YCLC6	138	0.0624 -		SFAM01.1	PPB	5/10/2024 12:57	SMPL
P2268-02	YCLC6D	182	0.072 -		SFAM01.1	PPB	5/10/2024 12:59	SMPL
P2268-03	YCLC6S	5266	1.1795 -		SFAM01.1	PPB	5/10/2024 13:01	SMPL
P2320-04	E1CZ3	-73	0.0165 -		SFAM01.1	PPB	5/10/2024 13:04	SMPL
P2320-05	E1D05	47	0.0426 -		SFAM01.1	PPB	5/10/2024 13:06	SMPL
CCV063	CCV063	20720	4.5459 -		SFAM01.1	PPB	5/10/2024 13:08	SMPL
CCB063	CCB063	-146	0.0006 -		SFAM01.1	PPB	5/10/2024 13:10	SMPL