

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130713

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.00	05/10/2024	12:50
2	CAL2	1	Water	NA	NA	20.3	7.01	05/10/2024	12:51
3	CAL3	1	Water	NA	NA	20.3	10.02	05/10/2024	12:55
4	ICV	1	Water	NA	NA	20.3	7.00	05/10/2024	13:00
5	CCV1	1	Water	NA	NA	20.2	2.01	05/10/2024	13:02
6	P2452-02	1	Water	NA	NA	20.7	5.96	05/10/2024	13:15
7	P2452-02DUP	1	Water	NA	NA	20.8	5.97	05/10/2024	13:19
8	P2452-03	1	Water	NA	NA	20.2	6.48	05/10/2024	13:25
9	CCV2	1	Water	NA	NA	20.2	12.01	05/10/2024	13:30

WB 30713

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p2452 WorkList ID : 180163 Department : Wet-Chemistry Date : 05-10-2024 12:31:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2452-02 C	TT-060-IDWGW-20240509	Water	pH	Cool 4 deg C	TETR06	L12	05/09/2024	9040C
P2452-03 C	TT-061-IDWGW-20240509	Water	pH	Cool 4 deg C	TETR06	L12	05/09/2024	9040C

Date/Time 05-10-24 12:30
Raw Sample Received by: JH SM
Raw Sample Relinquished by: JH SM

Date/Time 05-10-24 14:30
Raw Sample Received by: JH SM
Raw Sample Relinquished by: JH SM