


Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB130716
Analysis Date: 05/10/2024
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 05/10/2024
Extraction IN Time: 13:30
Extraction OUT Time: 13:37
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB130716BL	LB130716BL	WATER	1.3	1000	100	3.1563	3.1563	3.01	3.1564	3.1564	0.0001	0.1
2	LB130716BS	LB130716BS	WATER	1.3	1000	100	2.9600	2.9600	3.02	2.9771	2.9771	0.0171	17.1
3	LB130716BSD	LB130716BSD	WATER	1.3	1000	100	2.9774	2.9774	3.03	2.9949	2.9949	0.0175	17.5
4	P2403-07	LOD-MDL-WATER-01-QT2-2	WATER	1.3	1000	100	3.0177	3.0177	3.05	3.0198	3.0198	0.0021	2.1
5	P2403-08	LOQ-WATER-02-QT2-2024	WATER	1.3	1000	100	3.0854	3.0854	3.04	3.0896	3.0896	0.0042	4.2



Chemical Name	Chemical Lot #
HEXANE	W3076
pH Paper 0-14	NA
Sodium Sulfate	EP2479
1:1 HCL	WP106917
Silica Gel	NA
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP106918
LCSWD	5.00 ML	WP106919
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0019</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>15:31</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>14:30</u>		
Bal Check Time:	<u>13:55</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>16:10</u>
			Out Time1	<u>15:30</u>		

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>17:26</u>
1.0000 gram Balance:	<u>1.0004</u>	(0.9950-1.0050)	In Time2:	<u>16:40</u>		
Bal Check Time:	<u>18:10</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>18:00</u>
			Out Time2:	<u>17:25</u>		

WORKLIST(Hardcopy Internal Chain)

2730716

WorkList Name : non poplar p2403 WorkList ID : 180167 Department : Wet-Chemistry Date : 05-10-2024 12:57:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2403-07	LOD-MDL-WATER-01-QT2-202	Water	Non-Polar Material	Conc H2SO4 to pH < 2	CHEM02	QA Of	05/07/2024	1664A
P2403-08	LOQ-WATER-02-QT2-2024	Water	Non-Polar Material	Conc H2SO4 to pH < 2	CHEM02	QA Of	05/07/2024	1664A

Date/Time 05/10/24 13:20
Raw Sample Received by: JH GSC
Raw Sample Relinquished by: JH GSC

Date/Time 05/10/24 17:30
Raw Sample Received by: JH GSC
Raw Sample Relinquished by: JH GSC