

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130720

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	05/14/2024	13:10
2	P2418-01		1	14	100.00	>212.0	>212.0	05/14/2024	13:40
3	P2418-01DUP		1	14	100.00	>212.0	>212.0	05/14/2024	14:10
4	P2440-03		1	12	100.00	>212.0	>212.0	05/14/2024	14:40
5	P2454-02		1	12	100.00	>212.0	>212.0	05/14/2024	15:10
6	P2458-05		1	14	100.00	>212.0	>212.0	05/14/2024	15:40
7	P2458-06		1	12	100.00	>212.0	>212.0	05/14/2024	16:10

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

6130720

WorkList Name : fp-5-13

WorkList ID : 180195

Department : Wet-Chemistry

Date : 05-13-2024 09:52:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2418-01	1410	Water	Flash Point	Cool 4 deg C	PSEG03	K51	05/08/2024	1010B
P2440-03	EQ-TANK-CONTAINMENT-PIT	Water	Flash Point	Cool 4 deg C	PSEG03	L11	05/09/2024	1010B
P2454-02	SP25-06-20240510-WC	Water	Flash Point	Cool 4 deg C	PARS02	L21	05/10/2024	1010B
P2458-05	34790	Water	Flash Point	Cool 4 deg C	PSEG03	L21	05/10/2024	1010B
P2458-06	34789	Water	Flash Point	Cool 4 deg C	PSEG03	L21	05/10/2024	1010B

Date/Time 05/14/2024 13:00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: M. COC

Date/Time 05/14/2024 16:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC