

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130721

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	05/13/2024	08:15
2	P2437-06		1	14	100.00	>212.0	>212.0	05/13/2024	08:45
3	P2437-06DUP		1	14	100.00	>212.0	>212.0	05/13/2024	09:15
4	P2437-07		1	12	100.00	>212.0	>212.0	05/13/2024	09:45
5	P2437-08		1	13	100.00	>212.0	>212.0	05/13/2024	10:15
6	P2439-01		1	11	100.00	>212.0	>212.0	05/13/2024	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB130721

WorkList Name : fp-5-10-

WorkList ID : 180145

Department : Wet-Chemistry

Date : 05-10-2024 08:18:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2437-06	OUTSIDE-BLACK-DRUM	Water	Flash Point	Cool 4 deg C	PSEG03	K41	05/09/2024	1010B
P2437-07	OUTSIDE-RED-DRUM	Water	Flash Point	Cool 4 deg C	PSEG03	K41	05/09/2024	1010B
P2437-08	WAREHOUSE-TOTE-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	K41	05/09/2024	1010B
P2439-01	50319	Water	Flash Point	Cool 4 deg C	PSEG03	L11	05/09/2024	1010B

Date/Time 05/11/2024 08:10
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 05/11/2024 16:20
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC