

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130727

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/13/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	05/13/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	05/13/2024	08:12
4	ICV	1	Water	NA	NA	20.3	6.98	05/13/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	05/13/2024	08:20
6	P2454-02	1	Water	NA	NA	20.6	6.66	05/13/2024	08:32
7	P2454-02DUP	1	Water	NA	NA	20.7	6.67	05/13/2024	08:33
8	P2458-05	1	Water	NA	NA	20.1	5.76	05/13/2024	08:37
9	P2458-06	1	Water	NA	NA	20.7	7.64	05/13/2024	08:40
10	CCV2	1	Water	NA	NA	20.3	12.02	05/13/2024	08:45

WORKLIST(Hardcopy Internal Chain)

130727

WorkList Name : ph 2454

WorkList ID : 180171

Department : Wet-Chemistry

Date : 05-13-2024 07:55:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2454-02	C SP25-06-20240510-WC	Water	pH	Cool 4 deg C	PARS02	L21	05/10/2024	9040C
P2458-05	A 34790	Water	pH	Cool 4 deg C	PSEG03	L21	05/10/2024	9040C
P2458-06	A 34789	Water	pH	Cool 4 deg C	PSEG03	L21	05/10/2024	9040C

Date/Time 05.13.24 08:00

Raw Sample Received by: 70 (wr)

Raw Sample Relinquished by: [Signature]

Date/Time

05.13.24

Raw Sample Received by:

Raw Sample Relinquished by:

10:00

[Signature]

[Signature]