

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130728

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/13/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	05/13/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	05/13/2024	08:12
4	ICV	1	Water	NA	NA	20.3	6.98	05/13/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	05/13/2024	08:20
6	P2458-03	1	Solid	20.01	20	20.7	7.21	05/13/2024	08:55
7	P2458-04	1	Solid	20.03	20	20.7	6.78	05/13/2024	09:00
8	P2458-04DUP	1	Solid	20.04	20	20.8	7.65	05/13/2024	09:01
9	CCV2	1	Water	NA	NA	20.3	12.02	05/13/2024	09:05

WORKLIST(Hardcopy Internal Chain)

130328

WorkList Name : ph p2458 solids WorkList ID : 180172 Department : Wet-Chemistry Date : 05-13-2024 07:56:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2458-03	34786	Solid	pH	Cool 4 deg C	PSEG03	L21	05/10/2024	9045D
P2458-04	34733	Solid	pH	Cool 4 deg C	PSEG03	L21	05/10/2024	9045D

Date/Time 05-13-24 08:00
Raw Sample Received by: JWC
Raw Sample Relinquished by: JWC

Date/Time 05-13-24 10:30
Raw Sample Received by: JWC
Raw Sample Relinquished by: JWC