


LB130736

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9431e-004

C= -3.2692e-002

Rho= 0.9994537

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.003	0.003	183	0.000	183					90)
0.2	0.200	0.172	-0.028	1052	0.0 %	1052					-14
2.5	2.500	2.486	-0.014	12963	0.0 %	12963					-1
5.0	5.000	5.196	0.196	26910	0.0 %	26910					41
7.5	7.500	7.285	-0.215	37659	0.0 %	37659					-3
10.0	10.000	10.059	0.059	51935	0.0 %	51935					1

LB130736 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 May 2024 12:34:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	183	-	0.0000	7471B	PPB	13 May 2024 12:40:37	S	Std
0.2 - 1	502	1052	-	0.2000	7471B	PPB	13 May 2024 12:42:54	S	Std
2.5 - 1	52.5	12963	-	2.5000	7471B	PPB	13 May 2024 12:45:12	S	Std
5.0 - 1	55	26910	-	5.0000	7471B	PPB	13 May 2024 12:47:28	S	Std
7.5 - 1	57.5	37659	-	7.5000	7471B	PPB	13 May 2024 12:49:48	S	Std
10.0 - 1	510	51935	-	10.0000	7471B	PPB	13 May 2024 12:54:52	S	Std
ICV44 - 1	ICV44	21537	4.1521	-	7471B	PPB	13 May 2024 13:00:11	U	SMPL
ICB44 - 1	ICB44	-97	-0.0515	-	7471B	PPB	13 May 2024 13:02:28	U	SMPL
CCV69 - 1	CCV69	27874	5.3834	-	7471B	PPB	13 May 2024 13:04:46	U	SMPL
CCB69 - 1	CCB69	-62	-0.0447	-	7471B	PPB	13 May 2024 13:07:01	U	SMPL
CRA - 1	CRA	1104	0.1818	-	7471B	PPB	13 May 2024 13:09:21	U	SMPL
HighStd - 1	HighStd	52755	10.2179	-	7471B	PPB	13 May 2024 13:11:36	U	SMPL
ChkStd - 1	ChkStd	34529	6.6765	-	7471B	PPB	13 May 2024 13:13:52	U	SMPL
PB160878BL - 1	PBS	-238	-0.0789	-	7471B	PPB	13 May 2024 13:16:17	U	SMPL
PB160878BS - 1	LCSS	20311	3.9139	-	7471B	PPB	13 May 2024 13:18:40	U	SMPL
P2422-01 - 1	WCS-TP-D	13943	2.6765	-	7471B	PPB	13 May 2024 13:20:56	U	SMPL
P2422-05 - 1	WCS-TP-F	11796	2.2593	-	7471B	PPB	13 May 2024 13:23:16	U	SMPL
P2423-01 - 1	WC-GG	2455	0.4443	-	7471B	PPB	13 May 2024 13:25:37	U	SMPL
P2423-05 - 1	WC-HH	783	0.1194	-	7471B	PPB	13 May 2024 13:27:57	U	SMPL
P2430-01 - 1	1011UMR-SS001-1224-01	8802	1.6776	-	7471B	PPB	13 May 2024 13:30:13	U	SMPL
CCV70 - 1	CCV70	25695	4.9600	-	7471B	PPB	13 May 2024 13:32:29	U	SMPL
CCB70 - 1	CCB70	-128	-0.0576	-	7471B	PPB	13 May 2024 13:34:47	U	SMPL
P2430-03 - 1	1011UMR-SS002-1224-01	10389	1.9860	-	7471B	PPB	13 May 2024 13:37:07	U	SMPL
P2430-05 - 1	1011UMR-SS003-1224-01	8339	1.5876	-	7471B	PPB	13 May 2024 13:39:22	U	SMPL
P2433-01 - 1	72-12013	549	0.0740	-	7471B	PPB	13 May 2024 13:41:40	U	SMPL
P2437-01 - 1	359	483	0.0612	-	7471B	PPB	13 May 2024 13:43:59	U	SMPL
P2444-01 - 1	WC-DD	2892	0.5292	-	7471B	PPB	13 May 2024 13:46:15	U	SMPL
P2444-01DUP - 1	WC-DDDUP	2851	0.5213	-	7471B	PPB	13 May 2024 13:48:32	U	SMPL
P2444-01MS - 1	WC-DDMS	23856	4.6027	-	7471B	PPB	13 May 2024 13:50:49	U	SMPL
P2444-01MSD - 1	WC-DDMSD	23635	4.5597	-	7471B	PPB	13 May 2024 13:53:07	U	SMPL
P2444-01LX5 - 1		520	0.0683	-	7471B	PPB	13 May 2024 13:57:39	U	SMPL
P2444-01A - 1		23779	4.5877	-	7471B	PPB	13 May 2024 13:59:55	U	SMPL
CCV71 - 1	CCV71	28006	5.4090	-	7471B	PPB	13 May 2024 14:04:58	U	SMPL
CCB71 - 1	CCB71	-73	-0.0469	-	7471B	PPB	13 May 2024 14:07:13	U	SMPL