

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130749

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	05/14/2024	08:30
2	P2437-05		1	13	100.00	>212.0	>212.0	05/14/2024	09:00
3	P2437-05DUP		1	13	100.00	>212.0	>212.0	05/14/2024	09:30
4	P2440-01		1	11	100.00	>212.0	>212.0	05/14/2024	10:00
5	P2465-05		1	14	100.00	>212.0	>212.0	05/14/2024	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6130749

WorkList Name : FP-S-513

WorkList ID : 180210

Department : Wet-Chemistry

Date : 05-13-2024 16:34:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2437-05	WAREHOUSE-TOTE-OIL	Solid	Flash Point	Cool 4 deg C	PSEG03	K41	05/09/2024	1010B
P2440-01	DNALP-SLUDGE	Solid	Flash Point	Cool 4 deg C	PSEG03	L11	05/09/2024	1010B
P2465-05	SCRUBBER-LIQUID	Solid	Flash Point	Cool 4 deg C	PSEG03	K31	05/13/2024	1010B

Date/Time 05/16/2024 08:20
 Raw Sample Received by: RH CWC
 Raw Sample Relinquished by: JH CWC

Date/Time 05/16/2024 10:40
 Raw Sample Received by: JH CWC
 Raw Sample Relinquished by: RH CWC