

LB136754


7471B

/ INSTRUMENT ID: CV1

Linear

 μ Abs.:

53563



A= 0.0000e+000

B= 1.8674e-004

C= 5.3767e-002

Rho= 0.9998887

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.058	0.058	21	0.000	21					7.0
0.2	0.200	0.238	0.038	988	0.0 %	988					19
2.5	2.500	2.409	-0.091	12615	0.0 %	12615					-4
5.0	5.000	4.959	-0.041	26271	0.0 %	26271					-1
7.5	7.500	7.479	-0.021	39765	0.0 %	39765					0
10.0	10.000	10.056	0.056	53563	0.0 %	53563					1

LB130754 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 May 2024 13:11:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7471B	PPB	14 May 2024 13:15:46	S	Std
0.2 - 1	50.2	988	-	0.2000	7471B	PPB	14 May 2024 13:18:03	S	Std
2.5 - 1	52.5	12615	-	2.5000	7471B	PPB	14 May 2024 13:20:20	S	Std
5.0 - 1	55	26271	-	5.0000	7471B	PPB	14 May 2024 13:22:37	S	Std
7.5 - 1	57.5	39765	-	7.5000	7471B	PPB	14 May 2024 13:24:56	S	Std
10.0 - 1	51.0	53563	-	10.0000	7471B	PPB	14 May 2024 13:27:17	S	Std
ICV46 - 1	ICV46	20728	3.9244	-	7471B	PPB	14 May 2024 13:30:36	U	SMPL
ICB46 - 1	ICB46	-210	0.0146	-	7471B	PPB	14 May 2024 13:32:51	U	SMPL
CCV74 - 1	CCV74	26595	5.0200	-	7471B	PPB	14 May 2024 13:35:10	U	SMPL
CCB74 - 1	CCB74	-240	0.0090	-	7471B	PPB	14 May 2024 13:37:25	U	SMPL
CRA - 1	CRA	926	0.2267	-	7471B	PPB	14 May 2024 13:39:46	U	SMPL
HighStd - 1	HighStd	53603	10.0633	-	7471B	PPB	14 May 2024 13:42:01	U	SMPL
ChkStd - 1	ChkStd	36245	6.8220	-	7471B	PPB	14 May 2024 13:44:16	U	SMPL
PB160925BL - 1	PBS	-384	-0.0179	-	7471B	PPB	14 May 2024 13:46:42	U	SMPL
PB160925BS - 1	LCSS	22182	4.1959	-	7471B	PPB	14 May 2024 13:51:54	U	SMPL
P2465-05 - 1	SCRUBBER-LIQUID	1219	0.2814	-	7471B	PPB	14 May 2024 13:54:11	U	SMPL
P2466-01 - 1	OR-03-051324	5771	1.1314	-	7471B	PPB	14 May 2024 13:56:32	U	SMPL
P2467-01 - 1	HD-02-051324	3501	0.7075	-	7471B	PPB	14 May 2024 13:58:51	U	SMPL
P2468-01 - 1	WC-MH-II	1872	0.4033	-	7471B	PPB	14 May 2024 14:01:11	U	SMPL
P2469-01 - 1	MH-A	1118	0.2625	-	7471B	PPB	14 May 2024 14:03:28	U	SMPL
CCV75 - 1	CCV75	24878	4.6994	-	7471B	PPB	14 May 2024 14:05:45	U	SMPL
CCB75 - 1	CCB75	-448	-0.0299	-	7471B	PPB	14 May 2024 14:08:02	U	SMPL
P2470-01 - 1	WCS-TPG	35950	6.7669	-	7471B	PPB	14 May 2024 14:10:21	U	SMPL
P2470-01DUP - 1	WCS-TPGDUP	33997	6.4022	-	7471B	PPB	14 May 2024 14:12:36	U	SMPL
P2470-01MS - 1	WCS-TPGMS	60468	11.3453	-	7471B	PPB	14 May 2024 14:14:55	U	SMPL
P2470-01MSD - 1	WCS-TPGMSD	59435	11.1524	-	7471B	PPB	14 May 2024 14:17:20	U	SMPL
P2470-01LX5 - 1		5191	1.0231	-	7471B	PPB	14 May 2024 14:19:49	U	SMPL
P2470-01A - 1		58885	11.0497	-	7471B	PPB	14 May 2024 14:22:19	U	SMPL
CCV76 - 1	CCV76	24594	4.6463	-	7471B	PPB	14 May 2024 14:24:37	U	SMPL
CCB76 - 1	CCB76	-521	-0.0435	-	7471B	PPB	14 May 2024 14:27:16	U	SMPL
P2470-01DLX2 - 1		17434	3.3093	-	7471B	PPB	14 May 2024 14:32:11	U	SMPL
P2470-01DUPDLX2 - 1		15648	2.9758	-	7471B	PPB	14 May 2024 14:34:26	U	SMPL
P2470-01MSDLX2 - 1		26114	4.9302	-	7471B	PPB	14 May 2024 14:36:44	U	SMPL
P2470-01MSDDLX2 - 1		26267	4.9587	-	7471B	PPB	14 May 2024 14:39:02	U	SMPL
P2470-01LDLX10 - 1		2878	0.5912	-	7471B	PPB	14 May 2024 14:41:22	U	SMPL
P2470-01ADLX2 - 1		26352	4.9746	-	7471B	PPB	14 May 2024 14:43:44	U	SMPL
CCV77 - 1	CCV77	25346	4.7868	-	7471B	PPB	14 May 2024 14:46:01	U	SMPL
CCB77 - 1	CCB77	-595	-0.0573	-	7471B	PPB	14 May 2024 14:48:22	U	SMPL