

LB130762


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43385



A= 0.0000e+000

B= 2.2862e-004

C= 1.2320e-002

Rho= 0.9998494

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.025	0.025	57	0.000	57	0				1.031
0.2	0.200	0.207	0.007	853	0.0 %	853					1.031
2.5	2.500	2.414	-0.086	10505	0.0 %	10505					1.031
5.0	5.000	5.015	0.015	21884	0.0 %	21884					1.031
7.5	7.500	7.607	0.107	33221	0.0 %	33221					1.031
10.0	10.000	9.931	-0.069	43385	0.0 %	43385					1.031

LB130762 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 May 2024 08:22:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7470A	PPB	15 May 2024 08:42:42	S	Std
0.2 - 1	50.2	853	-	0.2000	7470A	PPB	15 May 2024 08:45:00	S	Std
2.5 - 1	52.5	10505	-	2.5000	7470A	PPB	15 May 2024 08:47:17	S	Std
5.0 - 1	55	21884	-	5.0000	7470A	PPB	15 May 2024 08:49:34	S	Std
7.5 - 1	57.5	33221	-	7.5000	7470A	PPB	15 May 2024 08:51:51	S	Std
10.0 - 1	510	43385	-	10.0000	7470A	PPB	15 May 2024 08:57:01	S	Std
ICV48 - 1	ICV48	17893	4.1030	-	7470A	PPB	15 May 2024 09:02:23	U	SMPL
ICB48 - 1	ICB48	-389	-0.0766	-	7470A	PPB	15 May 2024 09:04:40	U	SMPL
CCV81 - 1	CCV81	22430	5.1402	-	7470A	PPB	15 May 2024 09:06:57	U	SMPL
CCB81 - 1	CCB81	-39	0.0034	-	7470A	PPB	15 May 2024 09:11:50	U	SMPL
CRA - 1	CRA	756	0.1852	-	7470A	PPB	15 May 2024 09:14:05	U	SMPL
HighStd - 1	HighStd	43354	9.9238	-	7470A	PPB	15 May 2024 09:16:20	U	SMPL
ChkStd - 1	ChkStd	30586	7.0048	-	7470A	PPB	15 May 2024 09:18:36	U	SMPL
PB160895BL - 1	PBW	-64	-0.0023	-	7470A	PPB	15 May 2024 09:23:39	U	SMPL
PB160895BS - 1	LCSW	16678	3.8252	-	7470A	PPB	15 May 2024 09:26:55	U	SMPL
P2422-04 - 1	WCS-TP-D	-262	-0.0476	-	7470A	PPB	15 May 2024 09:29:11	U	SMPL
P2422-04DUP - 1	WCS-TP-DDUP	-18	0.0082	-	7470A	PPB	15 May 2024 09:31:30	U	SMPL
P2422-04MS - 1	WCS-TP-DMS	17854	4.0940	-	7470A	PPB	15 May 2024 09:33:47	U	SMPL
P2422-04MSD - 1	WCS-TP-DMSD	15552	3.5678	-	7470A	PPB	15 May 2024 09:36:04	U	SMPL
P2422-08 - 1	WCS-TP-F	-330	-0.0631	-	7470A	PPB	15 May 2024 09:38:22	U	SMPL
CCV82 - 1	CCV82	22916	5.2513	-	7470A	PPB	15 May 2024 09:40:40	U	SMPL
CCB82 - 1	CCB82	-263	-0.0478	-	7470A	PPB	15 May 2024 09:42:56	U	SMPL
P2423-04 - 1	WC-GG	-67	-0.0030	-	7470A	PPB	15 May 2024 09:45:13	U	SMPL
P2423-08 - 1	WC-HH	-65	-0.0025	-	7470A	PPB	15 May 2024 09:47:28	U	SMPL
P2430-02 - 1	1011UMR-SS001-1224-01	20	0.0169	-	7470A	PPB	15 May 2024 09:49:44	U	SMPL
P2430-04 - 1	1011UMR-SS002-1224-01	53	0.0244	-	7470A	PPB	15 May 2024 09:52:00	U	SMPL
P2430-06 - 1	1011UMR-SS003-1224-01	841	0.2046	-	7470A	PPB	15 May 2024 09:54:16	U	SMPL
P2437-03 - 1	SOM-44	139	0.0441	-	7470A	PPB	15 May 2024 09:56:34	U	SMPL
P2437-04 - 1	COMP-1	357	0.0939	-	7470A	PPB	15 May 2024 09:58:50	U	SMPL
P2437-05 - 1	WAREHOUSE-TOTE-OIL	430	0.1106	-	7470A	PPB	15 May 2024 10:01:07	U	SMPL
CCV83 - 1	CCV83	21445	4.9150	-	7470A	PPB	15 May 2024 10:06:03	U	SMPL
CCB83 - 1	CCB83	-38	0.0036	-	7470A	PPB	15 May 2024 10:11:12	U	SMPL
PB160833TB - 1	PB160833TB	-14	0.0091	-	7470A	PPB	15 May 2024 10:13:27	U	SMPL
P2422-04LX5 - 1		95	0.0340	-	7470A	PPB	15 May 2024 10:15:43	U	SMPL
P2422-04A - 1		15423	3.5383	-	7470A	PPB	15 May 2024 10:17:58	U	SMPL
CCV84 - 1	CCV84	22198	5.0871	-	7470A	PPB	15 May 2024 10:20:14	U	SMPL
CCB84 - 1	CCB84	-164	-0.0252	-	7470A	PPB	15 May 2024 10:25:13	U	SMPL