

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130765

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/15/2024	09:15
2	CAL2	1	Water	NA	NA	20.3	7.01	05/15/2024	09:16
3	CAL3	1	Water	NA	NA	20.3	10.02	05/15/2024	09:17
4	ICV	1	Water	NA	NA	20.3	7.00	05/15/2024	09:20
5	CCV1	1	Water	NA	NA	20.2	2.01	05/15/2024	09:25
6	P2476-02	1	Solid	20.02	20	22.2	6.70	05/15/2024	09:30
7	P2476-02DUP	1	Solid	20.03	20	22.3	6.71	05/15/2024	09:31
8	P2477-03	1	Solid	20.02	20	22.7	5.14	05/15/2024	09:37
9	P2477-04	1	Solid	20.02	20	23.0	5.49	05/15/2024	09:40
10	P2480-04	1	Solid	20.02	20	21.0	5.67	05/15/2024	09:45
11	CCV2	1	Water	NA	NA	20.3	12.02	05/15/2024	09:50

WORKLIST(Hardcopy Internal Chain)

UB 130765

WorkList Name : corrosivity p2480

WorkList ID : 180250

Department : Wet-Chemistry

Date : 05-15-2024 09:01:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2476-02	WC-1-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	K31	05/14/2024	9045D
P2477-03	ELIZ-24-00052	Solid	Corrosivity	Cool 4 deg C	PSEG03	K31	05/14/2024	9045D
P2477-04	COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	K31	05/14/2024	9045D
P2480-04	WC-MH-JJ	Solid	Corrosivity	Cool 4 deg C	PSEG03	K31	05/14/2024	9045D

Date/Time 05-15-24 09:10

Raw Sample Received by: 761 6011

Raw Sample Relinquished by: 761 54

Date/Time 05-15-24

Raw Sample Received by: 761 54

Raw Sample Relinquished by: 761 54