

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130766

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/15/2024	09:15
2	CAL2	1	Water	NA	NA	20.3	7.01	05/15/2024	09:16
3	CAL3	1	Water	NA	NA	20.3	10.02	05/15/2024	09:17
4	ICV	1	Water	NA	NA	20.3	7.00	05/15/2024	09:20
5	CCV1	1	Water	NA	NA	20.2	2.01	05/15/2024	09:25
6	P2476-01	1	Solid	20.02	20	22.2	6.70	05/15/2024	09:30
7	P2476-01DUP	1	Solid	20.03	20	22.3	6.71	05/15/2024	09:31
8	CCV2	1	Water	NA	NA	20.3	12.02	05/15/2024	09:50

UP 130760

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p2480 WorkList ID : 180249 Department : Wet-Chemistry Date : 05-15-2024 09:01:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2476-01	WC-1-COMP	Solid	pH	Cool 4 deg C	EART12	K31	05/14/2024	9045D

Date/Time 05-15-24 09:11:00
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC SH

Date/Time 05-15-24 12:00
Raw Sample Received by: JDC SH
Raw Sample Relinquished by: JDC C&C