


LB130776

7470A

INSTRUMENT ID: CU2

Linear



A= 0.0000e+000

B= 2.3872e-004

C= 4.4867e-002

Rho= 0.9992646

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	0.041	0.041	-16	0.000	-16	0				-
0.2	0.200	0.221	0.021	736	0.0 %	736					11
2.5	2.500	2.348	-0.152	9647	0.0 %	9647					-6
5.0	5.000	5.001	0.001	20761	0.0 %	20761					0
7.5	7.500	7.758	0.258	32310	0.0 %	32310					3
10.0	10.000	9.832	-0.168	40997	0.0 %	40997					-2

LB130776 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 May 2024 16:02:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-16	-	0.0000	7470A	PPB	15 May 2024 16:12:40	S	Std
0.2 - 1	50.2	736	-	0.2000	7470A	PPB	15 May 2024 16:14:57	S	Std
2.5 - 1	52.5	9647	-	2.5000	7470A	PPB	15 May 2024 16:17:14	S	Std
5.0 - 1	55	20761	-	5.0000	7470A	PPB	15 May 2024 16:19:31	S	Std
7.5 - 1	57.5	32310	-	7.5000	7470A	PPB	15 May 2024 16:21:49	S	Std
10.0 - 1	510	40997	-	10.0000	7470A	PPB	15 May 2024 16:24:07	S	Std
ICV49 - 1	ICV49	15131	3.6570	-	7470A	PPB	15 May 2024 16:26:57	U	SMPL
ICB49 - 1	ICB49	-307	-0.0284	-	7470A	PPB	15 May 2024 16:29:12	U	SMPL
CCV85 - 1	CCV85	21006	5.0594	-	7470A	PPB	15 May 2024 16:31:30	U	SMPL
CCB85 - 1	CCB85	-321	-0.0318	-	7470A	PPB	15 May 2024 16:33:45	U	SMPL
CRA - 1	CRA	710	0.2144	-	7470A	PPB	15 May 2024 16:36:02	U	SMPL
HighStd - 1	HighStd	42315	10.1463	-	7470A	PPB	15 May 2024 16:38:18	U	SMPL
ChkStd - 1	ChkStd	29647	7.1222	-	7470A	PPB	15 May 2024 16:40:34	U	SMPL
PB160939BL - 1	PBW	-490	-0.0721	-	7470A	PPB	15 May 2024 16:42:52	U	SMPL
PB160939BS - 1	LCSW	18278	4.4082	-	7470A	PPB	15 May 2024 16:45:11	U	SMPL
P2444-04 - 1	WC-DD	-317	-0.0308	-	7470A	PPB	15 May 2024 16:47:28	U	SMPL
P2468-04 - 1	WC-MH-II	-34	0.0368	-	7470A	PPB	15 May 2024 16:49:46	U	SMPL
P2469-04 - 1	MH-A	-1	0.0446	-	7470A	PPB	15 May 2024 16:52:03	U	SMPL
P2470-04 - 1	WCS-TPG	47	0.0561	-	7470A	PPB	15 May 2024 16:54:20	U	SMPL
P2470-04DUP - 1	WCS-TPGDUP	-19	0.0403	-	7470A	PPB	15 May 2024 16:56:36	U	SMPL
CCV86 - 1	CCV86	21111	5.0845	-	7470A	PPB	15 May 2024 16:58:51	U	SMPL
CCB86 - 1	CCB86	-264	-0.0182	-	7470A	PPB	15 May 2024 17:01:06	U	SMPL
P2470-04MS - 1	WCS-TPGMS	17883	4.3139	-	7470A	PPB	15 May 2024 17:03:25	U	SMPL
P2470-04MSD - 1	WCS-TPGMSD	17859	4.3082	-	7470A	PPB	15 May 2024 17:05:40	U	SMPL
P2470-04LX5 - 1		-227	-0.0093	-	7470A	PPB	15 May 2024 17:07:58	U	SMPL
P2470-04A - 1		17985	4.3383	-	7470A	PPB	15 May 2024 17:10:16	U	SMPL
PB160868TB - 1	PB160868TB	-285	-0.0232	-	7470A	PPB	15 May 2024 17:12:32	U	SMPL
CCV87 - 1	CCV87	21758	5.2390	-	7470A	PPB	15 May 2024 17:14:50	U	SMPL
CCB87 - 1	CCB87	-313	-0.0299	-	7470A	PPB	15 May 2024 17:17:07	U	SMPL