

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130785

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/16/2024	08:30
2	CAL2	1	Water	NA	NA	20.3	7.00	05/16/2024	08:31
3	CAL3	1	Water	NA	NA	20.3	10.03	05/16/2024	08:33
4	ICV	1	Water	NA	NA	20.3	7.02	05/16/2024	08:35
5	CCV1	1	Water	NA	NA	20.3	2.02	05/16/2024	08:40
6	P2491-03	1	Water	NA	NA	20.1	6.36	05/16/2024	08:43
7	P2491-04	1	Water	NA	NA	21.1	6.58	05/16/2024	08:45
8	P2491-05	1	Water	NA	NA	22.2	7.86	05/16/2024	08:49
9	P2493-04	1	Water	NA	NA	20.9	7.28	05/16/2024	08:52
10	P2493-04DUP	1	Water	NA	NA	20.8	7.29	05/16/2024	08:53
11	CCV2	1	Water	NA	NA	20.3	12.03	05/16/2024	08:54

WORKLIST(Hardcopy Internal Chain)

VB 130785

WorkList Name : ph p2491 WorkList ID : 180276 Department : Wet-Chemistry Date : 05-16-2024 08:15:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2491-03	9 2988	Water	pH	Cool 4 deg C	PSEG03	K31	05/15/2024	9040C
P2491-04	A 3015	Water	pH	Cool 4 deg C	PSEG03	K31	05/15/2024	9040C
P2491-05	A 3021	Water	pH	Cool 4 deg C	PSEG03	K31	05/15/2024	9040C
P2493-04	A T244-47	Water	pH	Cool 4 deg C	PSEG03	K21	05/15/2024	9040C

Date/Time 05-16-24 08:25
Raw Sample Received by: JD CDC /
Raw Sample Relinquished by: RH CWC

Date/Time 05-16-24 11:30
Raw Sample Received by: RH CWC
Raw Sample Relinquished by: JD CDC