

LB130786


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52873



A= 0.0000e+000

B= 1.8936e-004

C= -8.7093e-003

Rho= 0.9994328

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/O D
0.0	0.000	0.002	0.002	58	0.000	58					0/0 D
0.2	0.200	0.176	-0.024	973	0.0 %	973					-12
2.5	2.500	2.428	-0.072	12868	0.0 %	12868					-3
5.0	5.000	5.249	0.249	27766	0.0 %	27766					5
7.5	7.500	7.342	-0.158	38821	0.0 %	38821					-2
10.0	10.000	10.003	0.003	52873	0.0 %	52873					0

LB130786 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 May 2024 09:20:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.0000	7471B	PPB	16 May 2024 09:45:15	S	Std
0.2 - 1	50.2	973	-	0.2000	7471B	PPB	16 May 2024 09:47:32	S	Std
2.5 - 1	52.5	12868	-	2.5000	7471B	PPB	16 May 2024 09:49:49	S	Std
5.0 - 1	55	27766	-	5.0000	7471B	PPB	16 May 2024 09:52:06	S	Std
7.5 - 1	57.5	38821	-	7.5000	7471B	PPB	16 May 2024 09:54:25	S	Std
10.0 - 1	510	52873	-	10.0000	7471B	PPB	16 May 2024 09:56:47	S	Std
ICV50 - 1	ICV50	23034	4.3529	-	7471B	PPB	16 May 2024 10:02:18	U	SMPL
ICB50 - 1	ICB50	-269	-0.0596	-	7471B	PPB	16 May 2024 10:04:34	U	SMPL
CCV88 - 1	CCV88	28084	5.3091	-	7471B	PPB	16 May 2024 10:06:56	U	SMPL
CCB88 - 1	CCB88	-277	-0.0612	-	7471B	PPB	16 May 2024 10:09:11	U	SMPL
CRA - 1	CRA	937	0.1687	-	7471B	PPB	16 May 2024 10:11:33	U	SMPL
HighStd - 1	HighStd	52057	9.8486	-	7471B	PPB	16 May 2024 10:13:48	U	SMPL
ChkStd - 1	ChkStd	36269	6.8590	-	7471B	PPB	16 May 2024 10:16:04	U	SMPL
PB160943BL - 1	PBS	-350	-0.0750	-	7471B	PPB	16 May 2024 10:18:31	U	SMPL
PB160943BS - 1	LCSS	23399	4.4220	-	7471B	PPB	16 May 2024 10:20:57	U	SMPL
P2473-01 - 1	TR-04-051424	2352	0.4367	-	7471B	PPB	16 May 2024 10:27:46	U	SMPL
P2473-01DUP - 1	TR-04-051424DUP	1983	0.3668	-	7471B	PPB	16 May 2024 10:30:04	U	SMPL
P2473-01MS - 1	TR-04-051424MS	21635	4.0880	-	7471B	PPB	16 May 2024 10:32:23	U	SMPL
P2473-01MSD - 1	TR-04-051424MSD	25830	4.8823	-	7471B	PPB	16 May 2024 10:34:42	U	SMPL
P2474-01 - 1	B-114-SB02	29839	5.6415	-	7471B	PPB	16 May 2024 10:37:01	U	SMPL
CCV89 - 1	CCV89	25808	4.8782	-	7471B	PPB	16 May 2024 10:39:23	U	SMPL
CCB89 - 1	CCB89	-293	-0.0642	-	7471B	PPB	16 May 2024 10:41:47	U	SMPL
P2474-02 - 1	B-114-SB01	4107	0.7690	-	7471B	PPB	16 May 2024 10:44:07	U	SMPL
P2476-01 - 1	WC-1-COMP	3514	0.6567	-	7471B	PPB	16 May 2024 10:46:22	U	SMPL
P2477-01 - 1	ETGI-339	2306	0.4279	-	7471B	PPB	16 May 2024 10:49:42	U	SMPL
P2480-01 - 1	WC-MH-JJ	670	0.1182	-	7471B	PPB	16 May 2024 10:51:57	U	SMPL
P2481-01 - 1	INK	61	0.0028	-	7471B	PPB	16 May 2024 10:54:14	U	SMPL
P2486-01 - 1	WCS-TPK	670	0.1182	-	7471B	PPB	16 May 2024 10:56:30	U	SMPL
P2486-05 - 1	WCS-TPL	3095	0.5773	-	7471B	PPB	16 May 2024 10:58:47	U	SMPL
P2487-01 - 1	WC-A-COMP	11580	2.1840	-	7471B	PPB	16 May 2024 11:01:04	U	SMPL
P2488-01 - 1	SP-G-COMP	16270	3.0721	-	7471B	PPB	16 May 2024 11:03:22	U	SMPL
P2490-01 - 1	72-11978	6587	1.2386	-	7471B	PPB	16 May 2024 11:05:41	U	SMPL
CCV90 - 1	CCV90	27952	5.2842	-	7471B	PPB	16 May 2024 11:08:00	U	SMPL
CCB90 - 1	CCB90	-189	-0.0445	-	7471B	PPB	16 May 2024 11:10:18	U	SMPL
P2491-01 - 1	VNJ-277	5445	1.0223	-	7471B	PPB	16 May 2024 11:12:40	U	SMPL
P2496-01 - 1	CHRT-24376	17342	3.2751	-	7471B	PPB	16 May 2024 11:14:56	U	SMPL
P2496-03 - 1	CHRT-28452	4044	0.7570	-	7471B	PPB	16 May 2024 11:17:13	U	SMPL
P2497-01 - 1	280856	1496	0.2746	-	7471B	PPB	16 May 2024 11:19:32	U	SMPL
P2497-03 - 1	RT3823	2047	0.3789	-	7471B	PPB	16 May 2024 11:21:50	U	SMPL
P2497-05 - 1	RT3997	5040	0.9456	-	7471B	PPB	16 May 2024 11:24:06	U	SMPL
P2497-07 - 1	280646	1079	0.1956	-	7471B	PPB	16 May 2024 11:26:24	U	SMPL
P2473-01LX5 - 1		286	0.0454	-	7471B	PPB	16 May 2024 11:31:33	U	SMPL
P2473-01A - 1		24707	4.6697	-	7471B	PPB	16 May 2024 11:33:49	U	SMPL
CCV91 - 1	CCV91	28232	5.3372	-	7471B	PPB	16 May 2024 11:36:05	U	SMPL
CCB91 - 1	CCB91	-162	-0.0394	-	7471B	PPB	16 May 2024 11:38:25	U	SMPL