

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130790

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	05/16/2024	11:10
2	P2493-02		1	4	8.00	46.4	46.7	05/16/2024	11:40
3	P2493-03		1	4	7.00	44.6	44.9	05/16/2024	12:40
4	P2501-01		1	11	100.00	>212.0	>212.0	05/16/2024	13:10
5	P2501-01DUP		1	11	100.00	>212.0	>212.0	05/16/2024	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66130790

WorkList Name : fp-s5-16

WorkList ID : 180290

Department : Wet-Chemistry

Date : 05-16-2024 09:00:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2493-02 A	D241	Solid	Flash Point	Cool 4 deg C	PSEG03	K21	05/15/2024	1010B
P2493-03 B	D242	Solid	Flash Point	Cool 4 deg C	PSEG03	K21	05/15/2024	1010B
P2501-01 A	33351	Solid	Flash Point	Cool 4 deg C	PSEG03	K31	05/16/2024	1010B

Date/Time 05/16/2024 14:00
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: JH GDC

Date/Time 05/16/2024 14:15
Raw Sample Received by: JH GDC
Raw Sample Relinquished by: RM CWC