

LB130792


7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

39684



A= 0.0000e+000

B= 2.4710e-004

C= 2.2587e-002

Rho= 0.9986034

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.0	0.000	0.036	0.036	54	0.000	54	0				11
0.2	0.200	0.222	0.022	805	0.0 %	805					0
2.5	2.500	2.501	0.001	10028	0.0 %	10028					-5
5.0	5.000	4.748	-0.252	19124	0.0 %	19124					5
7.5	7.500	7.865	0.365	31739	0.0 %	31739					-2
10.0	10.000	9.829	-0.171	39684	0.0 %	39684					

LB130792 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 May 2024 13:41:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7470A	PPB	16 May 2024 13:48:15	S	Std
0.2 - 1	50.2	805	-	0.2000	7470A	PPB	16 May 2024 13:50:32	S	Std
2.5 - 1	52.5	10028	-	2.5000	7470A	PPB	16 May 2024 13:57:43	S	Std
5.0 - 1	55	19124	-	5.0000	7470A	PPB	16 May 2024 14:00:00	S	Std
7.5 - 1	57.5	31739	-	7.5000	7470A	PPB	16 May 2024 14:07:37	S	Std
10.0 - 1	510	39684	-	10.0000	7470A	PPB	16 May 2024 14:15:28	S	Std
ICV53 - 1	ICV53	14875	3.6982	-	7470A	PPB	16 May 2024 14:18:19	U	SMPL
ICB53 - 1	ICB53	-321	-0.0567	-	7470A	PPB	16 May 2024 14:20:34	U	SMPL
CCV97 - 1	CCV97	19956	4.9537	-	7470A	PPB	16 May 2024 14:22:52	U	SMPL
CCB97 - 1	CCB97	-313	-0.0548	-	7470A	PPB	16 May 2024 14:25:07	U	SMPL
CRA - 1	CRA	722	0.2010	-	7470A	PPB	16 May 2024 14:27:24	U	SMPL
HighStd - 1	HighStd	39406	9.7599	-	7470A	PPB	16 May 2024 14:29:39	U	SMPL
ChkStd - 1	ChkStd	28362	7.0309	-	7470A	PPB	16 May 2024 14:31:55	U	SMPL
PB160953BL - 1	PBW	-470	-0.0936	-	7470A	PPB	16 May 2024 14:34:13	U	SMPL
PB160953BS - 1	LCSW	14694	3.6535	-	7470A	PPB	16 May 2024 14:36:32	U	SMPL
P2437-06 - 1	OUTSIDE-BLACK-DRUM	-239	-0.0365	-	7470A	PPB	16 May 2024 14:38:48	U	SMPL
P2437-06DUP - 1	OUTSIDE-BLACK-DRUMDUP	-2	0.0221	-	7470A	PPB	16 May 2024 14:41:07	U	SMPL
P2437-06MS - 1	OUTSIDE-BLACK-DRUMMS	13812	3.4356	-	7470A	PPB	16 May 2024 14:43:24	U	SMPL
P2437-06MSD - 1	OUTSIDE-BLACK-DRUMMSD	13718	3.4123	-	7470A	PPB	16 May 2024 14:45:41	U	SMPL
P2437-07 - 1	OUTSIDE-RED-DRUM	-209	-0.0291	-	7470A	PPB	16 May 2024 14:47:58	U	SMPL
CCV98 - 1	CCV98	19906	4.9414	-	7470A	PPB	16 May 2024 14:50:15	U	SMPL
CCB98 - 1	CCB98	-276	-0.0456	-	7470A	PPB	16 May 2024 14:52:30	U	SMPL
P2454-02 - 1	SP25-06-20240510-WC	-55	0.0090	-	7470A	PPB	16 May 2024 14:54:48	U	SMPL
PB160836TB - 1	PB160836TB	1	0.0228	-	7470A	PPB	16 May 2024 14:57:03	U	SMPL
PB160863TB - 1	PB160863TB	-8	0.0206	-	7470A	PPB	16 May 2024 14:59:18	U	SMPL
P2437-06LX5 - 1		107	0.0490	-	7470A	PPB	16 May 2024 15:01:34	U	SMPL
P2437-06A - 1		15234	3.7869	-	7470A	PPB	16 May 2024 15:03:50	U	SMPL
CCV99 - 1	CCV99	19579	4.8606	-	7470A	PPB	16 May 2024 15:06:06	U	SMPL
CCB99 - 1	CCB99	-307	-0.0533	-	7470A	PPB	16 May 2024 15:08:24	U	SMPL