

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130800

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/16/2024	13:37
2	CAL2	1	Water	NA	NA	20.3	7.00	05/16/2024	13:39
3	CAL3	1	Water	NA	NA	20.3	10.02	05/16/2024	13:40
4	ICV	1	Water	NA	NA	20.3	7.01	05/16/2024	13:41
5	CCV1	1	Water	NA	NA	20.3	2.02	05/16/2024	13:45
6	P2501-03	1	Water	NA	NA	20.6	6.90	05/16/2024	13:50
7	P2501-03DUP	1	Water	NA	NA	20.7	6.91	05/16/2024	13:51
8	P2501-05	1	Water	NA	NA	20.7	6.66	05/16/2024	14:00
9	P2511-07	1	Water	NA	NA	20.1	6.98	05/16/2024	14:10
10	CCV2	1	Water	NA	NA	20.3	12.02	05/16/2024	14:15

WORKLIST(Hardcopy Internal Chain)

130800

WorkList Name : ph p2511

WorkList ID : 180292

Department : Wet-Chemistry

Date : 05-16-2024 13:23:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2501-03	50334	Water	pH	Cool 4 deg C	PSEG03	K31	05/16/2024	9040C
P2501-05	50332-50333-COMP	Water	pH	Cool 4 deg C	PSEG03	K31	05/16/2024	9040C
P2511-07	510-AB	Water	pH	Cool 4 deg C	PSEG03	K21	05/16/2024	9040C

Date/Time 05-16-24 13:30

Raw Sample Received by: JRC

Raw Sample Relinquished by: JRC

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: