


LB130809

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0617e-004

C= -4.5761e-003

Rho= 0.9988838

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	0.047	0.047	251	0.000	251	0				1
0.2	0.200	0.240	0.040	1187	0.0 %	1187					20
2.5	2.500	2.228	-0.272	10830	0.0 %	10830					11
5.0	5.000	5.170	0.170	25100	0.0 %	25100					3
7.5	7.500	7.701	0.201	37376	0.0 %	37376					3
10.0	10.000	9.813	-0.187	47617	0.0 %	47617					2

LB130807 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 May 2024 10:12:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	251	-	0.0000	7470A	PPB	17 May 2024 10:42:46	S	Std
0.2 - 1	50.2	1187	-	0.2000	7470A	PPB	17 May 2024 10:45:03	S	Std
2.5 - 1	52.5	10830	-	2.5000	7470A	PPB	17 May 2024 10:47:20	S	Std
5.0 - 1	55	25100	-	5.0000	7470A	PPB	17 May 2024 10:52:15	S	Std
7.5 - 1	57.5	37376	-	7.5000	7470A	PPB	17 May 2024 10:54:31	S	Std
10.0 - 1	510	47617	-	10.0000	7470A	PPB	17 May 2024 10:59:36	S	Std
ICV54 - 1	ICV54	21362	4.3997	-	7470A	PPB	17 May 2024 11:18:01	U	SMPL
ICB54 - 1	ICB54	-93	-0.0238	-	7470A	PPB	17 May 2024 11:20:16	U	SMPL
CCV01 - 1	CCV01	24015	4.9467	-	7470A	PPB	17 May 2024 11:22:33	U	SMPL
CCB01 - 1	CCB01	-62	-0.0174	-	7470A	PPB	17 May 2024 11:24:48	U	SMPL
CRA - 1	CRA	1063	0.2146	-	7470A	PPB	17 May 2024 11:27:06	U	SMPL
HighStd - 1	HighStd	48191	9.9311	-	7470A	PPB	17 May 2024 11:29:21	U	SMPL
ChkStd - 1	ChkStd	35893	7.3956	-	7470A	PPB	17 May 2024 11:31:37	U	SMPL
PB160963BL - 1	PBW	-151	-0.0357	-	7470A	PPB	17 May 2024 11:33:55	U	SMPL
PB160963BS - 1	LCSW	18208	3.7494	-	7470A	PPB	17 May 2024 11:36:13	U	SMPL
P2476-02 - 1	WC-1-COMP	-252	-0.0565	-	7470A	PPB	17 May 2024 11:38:30	U	SMPL
P2477-03 - 1	ELIZ-24-00052	196	0.0358	-	7470A	PPB	17 May 2024 11:40:48	U	SMPL
P2477-04 - 1	COMP-1	53	0.0064	-	7470A	PPB	17 May 2024 11:43:05	U	SMPL
P2480-04 - 1	WC-MH-JJ	89	0.0138	-	7470A	PPB	17 May 2024 11:45:22	U	SMPL
P2486-04 - 1	WCS-TPK	341	0.0657	-	7470A	PPB	17 May 2024 11:47:37	U	SMPL
CCV02 - 1	CCV02	24495	5.0456	-	7470A	PPB	17 May 2024 11:49:53	U	SMPL
CCB02 - 1	CCB02	-15	-0.0077	-	7470A	PPB	17 May 2024 11:52:09	U	SMPL
P2486-08 - 1	WCS-TPL	35	0.0026	-	7470A	PPB	17 May 2024 11:54:26	U	SMPL
P2487-02 - 1	WC-A-COMP	115	0.0191	-	7470A	PPB	17 May 2024 11:56:41	U	SMPL
P2488-02 - 1	SP-G-COMP	102	0.0165	-	7470A	PPB	17 May 2024 11:58:57	U	SMPL
P2493-01 - 1	D243	123	0.0208	-	7470A	PPB	17 May 2024 12:01:13	U	SMPL
P2493-02 - 1	D241	781	0.1564	-	7470A	PPB	17 May 2024 12:03:29	U	SMPL
P2493-03 - 1	D242	529	0.1045	-	7470A	PPB	17 May 2024 12:05:45	U	SMPL
P2497-02 - 1	280856	-110	-0.0273	-	7470A	PPB	17 May 2024 12:08:02	U	SMPL
P2497-04 - 1	RT3823	-1	-0.0048	-	7470A	PPB	17 May 2024 12:10:19	U	SMPL
P2497-06 - 1	RT3997	-49	-0.0147	-	7470A	PPB	17 May 2024 12:12:36	U	SMPL
P2497-08 - 1	280646	17	-0.0011	-	7470A	PPB	17 May 2024 12:14:52	U	SMPL
CCV03 - 1	CCV03	22867	4.7100	-	7470A	PPB	17 May 2024 12:19:56	U	SMPL
CCB03 - 1	CCB03	-157	-0.0369	-	7470A	PPB	17 May 2024 12:22:11	U	SMPL
P2497-08DUP - 1	280646DUP	288	0.0548	-	7470A	PPB	17 May 2024 12:24:29	U	SMPL
P2497-08MS - 1	280646MS	18005	3.7076	-	7470A	PPB	17 May 2024 12:26:46	U	SMPL
P2497-08MSD - 1	280646MSD	18283	3.7649	-	7470A	PPB	17 May 2024 12:29:01	U	SMPL
PB160964BL - 1	PBW	-326	-0.0718	-	7470A	PPB	17 May 2024 12:31:19	U	SMPL
PB160964BS - 1	LCSW	19065	3.9261	-	7470A	PPB	17 May 2024 12:33:37	U	SMPL
P2491-03 - 1	2988	-262	-0.0586	-	7470A	PPB	17 May 2024 12:35:54	U	SMPL
P2491-04 - 1	3015	305	0.0583	-	7470A	PPB	17 May 2024 12:38:12	U	SMPL
P2491-05 - 1	3021	65	0.0088	-	7470A	PPB	17 May 2024 12:40:29	U	SMPL
P2493-04 - 1	T244-47	189	0.0344	-	7470A	PPB	17 May 2024 12:42:46	U	SMPL
P2493-04DUP - 1	T244-47DUP	49	0.0055	-	7470A	PPB	17 May 2024 12:45:02	U	SMPL
CCV04 - 1	CCV04	25320	5.2157	-	7470A	PPB	17 May 2024 12:47:17	U	SMPL
CCB04 - 1	CCB04	-233	-0.0526	-	7470A	PPB	17 May 2024 12:49:33	U	SMPL
P2493-04MS - 1	T244-47MS	20287	4.1780	-	7470A	PPB	17 May 2024 12:51:51	U	SMPL
P2493-04MSD - 1	T244-47MSD	19247	3.9636	-	7470A	PPB	17 May 2024 12:54:07	U	SMPL
PB160935TB - 1	PB160935TB	-253	-0.0567	-	7470A	PPB	17 May 2024 12:56:25	U	SMPL
PB160944TB - 1	PB160944TB	269	0.0509	-	7470A	PPB	17 May 2024 12:58:44	U	SMPL
P2497-08LX5 - 1		108	0.0177	-	7470A	PPB	17 May 2024 13:01:02	U	SMPL
P2497-08A - 1		17731	3.6511	-	7470A	PPB	17 May 2024 13:03:18	U	SMPL
P2493-04LX5 - 1		-61	-0.0172	-	7470A	PPB	17 May 2024 13:05:35	U	SMPL
P2493-04A - 1		19084	3.9300	-	7470A	PPB	17 May 2024 13:07:54	U	SMPL
CCV05 - 1	CCV05	25333	5.2184	-	7470A	PPB	17 May 2024 13:10:12	U	SMPL
CCB05 - 1	CCB05	-282	-0.0627	-	7470A	PPB	17 May 2024 13:12:30	U	SMPL