

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130823

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/17/2024	09:10
2	P2491-03		1	12	100.00	>212.0	>212.0	05/17/2024	10:45
3	P2491-03DUP		1	12	100.00	>212.0	>212.0	05/17/2024	11:15
4	P2491-04		1	14	100.00	>212.0	>212.0	05/17/2024	12:45
5	P2491-05		1	11	100.00	>212.0	>212.0	05/17/2024	13:15
6	P2493-04		1	11	100.00	>212.0	>212.0	05/17/2024	14:10
7	P2501-03		1	14	100.00	>212.0	>212.0	05/17/2024	14:40
8	P2501-05		1	11	100.00	>212.0	>212.0	05/17/2024	15:10
9	P2511-07		1	12	100.00	>212.0	>212.0	05/17/2024	15:40
10	P2511-09		1	14	100.00	>212.0	>212.0	05/17/2024	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6130823

WorkList Name : fp5-16

WorkList ID : 180280

Department : Wet-Chemistry

Date : 05-16-2024 08:27:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2491-03	2988	Water	Flash Point	Cool 4 deg C	PSEG03	K31	05/15/2024	1010B
P2491-04	3015	Water	Flash Point	Cool 4 deg C	PSEG03	K31	05/15/2024	1010B
P2491-05	3021	Water	Flash Point	Cool 4 deg C	PSEG03	K31	05/15/2024	1010B
P2493-04	T244-47	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/15/2024	1010B
P2501-03	50334	Water	Flash Point	Cool 4 deg C	PSEG03	K31	05/16/2024	1010B
P2501-05	50332-50333-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	K31	05/16/2024	1010B
P2511-07	510-AB	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/16/2024	1010B
P2511-09	1598	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/16/2024	1010B

Date/Time 05/17/2024 09:00
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: 12(20)

Date/Time 05/17/2024 16:15
 Raw Sample Received by: 12(20)
 Raw Sample Relinquished by: RM CWL