

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB130835

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP108023
Turbidity Calibration std, 20NTU	WP108028
Turbidity Calibration std, 40NTU	WP108024
Turbidity Calibration std, 1NTU	WP108025
Turbidity Calibration std, 5NTU	WP108026
Turbidity Calibration - CCV std, 10 NTU	WP108027
10 NTU Standard 500 ml	W2971

Intercept: -0.0592

Slope: 0.955

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.22		05/20/2024	13:30
2	CAL2	1	1	1.110	1.22	22.4	05/20/2024	13:34
3	CAL3	5	1	4.390	4.66	-6.8	05/20/2024	13:38
4	CAL4	10	1	9.480	9.99	-0.1	05/20/2024	13:42
5	CAL5	20	1	18.800	19.75	-1.3	05/20/2024	13:46
6	CAL6	40	1	38.300	40.17	0.4	05/20/2024	13:50



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB130835

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.400	9.905	05/20/2024	13:54
2	ICB		1	0.142	0.211	05/20/2024	13:58
3	CCV1	10	1	9.400	9.905	05/20/2024	14:02
4	CCB1		1	0.140	0.209	05/20/2024	14:06
5	LB130835BL		1	0.150	0.219	05/20/2024	14:10
6	P2403-07		1	0.520	0.606	05/20/2024	14:14
7	P2403-08		1	1.120	1.235	05/20/2024	14:18
8	CCV2	10	1	9.400	9.905	05/20/2024	14:22
9	CCB2		1	0.148	0.217	05/20/2024	14:26