

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130848

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/21/2024	09:30
2	CAL2	1	Water	NA	NA	20.3	7.00	05/21/2024	09:31
3	CAL3	1	Water	NA	NA	20.3	10.02	05/21/2024	09:35
4	ICV	1	Water	NA	NA	20.3	7.00	05/21/2024	09:40
5	CCV1	1	Water	NA	NA	20.3	2.01	05/21/2024	09:44
6	P2535-01	1	Water	NA	NA	23.3	6.23	05/21/2024	09:48
7	P2536-01	1	Water	NA	NA	23.6	6.09	05/21/2024	09:52
8	P2538-04	1	Water	NA	NA	22.4	7.41	05/21/2024	09:55
9	P2541-01	1	Water	NA	NA	24.4	11.69	05/21/2024	10:05
10	P2541-01DUP	1	Water	NA	NA	24.5	11.70	05/21/2024	10:06
11	CCV2	1	Water	NA	NA	20.3	12.02	05/21/2024	10:15

WORKLIST(Hardcopy Internal Chain)

UP 190848

WorkList Name : PH P2541

WorkList ID : 180386

Department : Wet-Chemistry

Date : 05-21-2024 09:20:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2535-01	MUL-OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	K21	05/20/2024	9040C
P2536-01	EJER-OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	L41	05/20/2024	9040C
P2538-04	A852	Water	pH	Cool 4 deg C	PSEG03	K31	05/20/2024	9040C
P2541-01	50322	Water	pH	Cool 4 deg C	PSEG03	K31	05/20/2024	9040C

Date/Time 05/21/24 09:25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 05/21/24 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]