

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130863

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/21/2024	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/21/2024	16:16
3	CAL3	1	Water	NA	NA	20.3	10.02	05/21/2024	16:19
4	ICV	1	Water	NA	NA	20.3	7.01	05/21/2024	16:22
5	CCV1	1	Water	NA	NA	20.5	2.02	05/21/2024	16:25
6	P2563-01	1	Solid	20.02	20	24.6	6.96	05/21/2024	16:35
7	P2563-01DUP	1	Solid	20.03	20	24.7	6.97	05/21/2024	16:36
8	CCV2	1	Water	NA	NA	20.3	12.02	05/21/2024	16:40

WORKLIST(Hardcopy Internal Chain)

VP 130863

WorkList Name : PH P2563

WorkList ID : 180413

Department : Wet-Chemistry

Date : 05-21-2024 16:06:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2563-01	72-11985	Solid	pH	Cool 4 deg C	PSEG03	K21	05/21/2024	9045D

Date/Time 05/21/24 16:10
Raw Sample Received by: JH GOC
Raw Sample Relinquished by: JH GOC

Date/Time 05/21/24 18:00
Raw Sample Received by: JH GOC
Raw Sample Relinquished by: JH GOC