

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130865

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/21/2024	15:10
2	CAL2	1	Water	NA	NA	20.3	7.00	05/21/2024	15:11
3	CAL3	1	Water	NA	NA	20.4	10.03	05/21/2024	15:15
4	ICV	1	Water	NA	NA	20.3	7.01	05/21/2024	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	05/21/2024	15:25
6	P2554-01	1	Water	NA	NA	24.6	6.75	05/21/2024	15:35
7	P2554-03	1	Water	NA	NA	24.6	5.39	05/21/2024	15:35
8	P2563-03	1	Water	NA	NA	24.2	6.82	05/21/2024	15:40
9	P2563-03DUP	1	Water	NA	NA	24.3	6.83	05/21/2024	15:41
10	CCV2	1	Water	NA	NA	20.3	12.02	05/21/2024	15:45

WORKLIST(Hardcopy Internal Chain)

W3 130815

WorkList Name : PH P2564 WorkList ID : 180403 Department : Wet-Chemistry Date : 05-21-2024 14:48:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2554-01 C	AUD-1041	Water	pH	Cool 4 deg C	PSEG03	K21	05/21/2024	9040C
P2554-03 C	AUD-1349	Water	pH	Cool 4 deg C	PSEG03	K21	05/21/2024	9040C
P2563-03 P	OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	K21	05/21/2024	9040C

Date/Time 05/21/23 15:00
Raw Sample Received by: 706001
Raw Sample Relinquished by: CP 5m

Date/Time 05/21/23 18:00
Raw Sample Received by: CP 5m
Raw Sample Relinquished by: 706001