

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130866

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	05/21/2024	10:40
2	P2535-01		1	12	100.00	>212.0	>212.0	05/21/2024	11:10
3	P2536-01		1	14	100.00	>212.0	>212.0	05/21/2024	11:40
4	P2537-01		1	12	100.00	>212.0	>212.0	05/21/2024	12:10
5	P2538-04		1	12	100.00	>212.0	>212.0	05/21/2024	12:40
6	P2541-01		1	14	100.00	>212.0	>212.0	05/21/2024	13:10
7	P2541-01DUP		1	14	100.00	>212.0	>212.0	05/21/2024	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB130866

WorkList Name : FP-5-20

WorkList ID : 180365

Department : Wet-Chemistry

Date : 05-20-2024 16:24:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2535-01D	MUL-OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/20/2024	1010B
P2536-01D	EJER-OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L41	05/20/2024	1010B
P2537-01B	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	K41	05/20/2024	1010B
P2538-04 Q	A852	Water	Flash Point	Cool 4 deg C	PSEG03	K31	05/20/2024	1010B
P2541-01F	50322	Water	Flash Point	Cool 4 deg C	PSEG03	K31	05/20/2024	1010B

Date/Time 05/21/2024 10:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JH CWC

Date/Time 05/21/2024 14:05
 Raw Sample Received by: JH CWC
 Raw Sample Relinquished by: RM CWC