

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB130871

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3085

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/22/2024	09:20
2	P2554-01		1	12	100.00	>212.0	>212.0	05/22/2024	09:50
3	P2554-01DUP		1	12	100.00	>212.0	>212.0	05/22/2024	10:20
4	P2554-03		1	13	100.00	>212.0	>212.0	05/22/2024	10:50
5	P2563-03		1	12	100.00	>212.0	>212.0	05/22/2024	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6130871

WorkList Name : fp522

WorkList ID : 180439

Department : Wet-Chemistry

Date : 05-22-2024 09:09:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2554-01	AUD-1041	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/21/2024	1010B
P2554-03	AUD-1349	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/21/2024	1010B
P2563-03	OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	K21	05/21/2024	1010B

Date/Time 05/22/2024 09:15
Raw Sample Received by: 12(sj)
Raw Sample Relinquished by: 706001

Date/Time 05/22/2024 11:55
Raw Sample Received by: M cooley
Raw Sample Relinquished by: 12(sj)