

LB130896


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47289



A= 0.0000e+000

B= 2.1368e-004 slope

C= 1.1025e-002 μ -intercept

Rho= 0.9997936

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/6/11
0.00	0.000	0.035	0.035	111	0.000	111					=
0.05	0.050					442	0				13
0.20	0.200	0.226	0.026	1006	0.0 %	1006					0
2.50	2.500	2.508	0.008	11685	0.0 %	11685					1
5.00	5.000	4.933	-0.067	23036	0.0 %	23036					2
7.50	7.500	7.382	-0.118	34495	0.0 %	34495					1
10.0	10.000	10.116	0.116	47289	0.0 %	47289					

INSTRUMENT ID ; CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	111	-		0 SFAM01.1	PPB	5/23/2024 11:21	Std
	0.2 S01	1006	-		0.2 SFAM01.1	PPB	5/23/2024 11:23	Std
	2.5 S02	11685	-		2.5 SFAM01.1	PPB	5/23/2024 11:26	Std
	5 S03	23036	-		5 SFAM01.1	PPB	5/23/2024 11:28	Std
	7.5 S04	34495	-		7.5 SFAM01.1	PPB	5/23/2024 11:30	Std
	10 S05	47289	-		10 SFAM01.1	PPB	5/23/2024 11:37	Std
ICV061	ICV061	18695	4.0058 -		SFAM01.1	PPB	5/23/2024 11:40	SMPL
ICB061	ICB061	-67	-0.0033 -		SFAM01.1	PPB	5/23/2024 11:42	SMPL
CCV033	CCV033	23783	5.0931 -		SFAM01.1	PPB	5/23/2024 11:45	SMPL
CCB033	CCB033	-108	-0.0121 -		SFAM01.1	PPB	5/23/2024 11:47	SMPL
PB161090BL	PBS090	92	0.0307 -		SFAM01.1	PPB	5/23/2024 11:49	SMPL
P2550-01	MBH5N9	6242	1.3448 -		SFAM01.1	PPB	5/23/2024 11:52	SMPL
P2550-02	MBH5P0	9111	1.9579 -		SFAM01.1	PPB	5/23/2024 11:54	SMPL
P2550-03	MBH5P1	7586	1.632 -		SFAM01.1	PPB	5/23/2024 11:56	SMPL
P2550-04	MBH5P2	4938	1.0662 -		SFAM01.1	PPB	5/23/2024 11:58	SMPL
P2550-05	MBH5P3	2487	0.5425 -		SFAM01.1	PPB	5/23/2024 12:01	SMPL
P2550-06	MBH5W8	6335	1.3647 -		SFAM01.1	PPB	5/23/2024 12:03	SMPL
P2550-07	MBH5W9	6542	1.4089 -		SFAM01.1	PPB	5/23/2024 12:05	SMPL
P2550-08	MBH5X0	4420	0.9555 -		SFAM01.1	PPB	5/23/2024 12:08	SMPL
P2550-09	MBH5X1	4637	1.0019 -		SFAM01.1	PPB	5/23/2024 12:10	SMPL
P2550-10	MBH5X2	1244	0.2768 -		SFAM01.1	PPB	5/23/2024 12:12	SMPL
P2550-11	MBH5X3	15471	3.3169 -		SFAM01.1	PPB	5/23/2024 12:14	SMPL
P2550-12	MBH5X4	17215	3.6896 -		SFAM01.1	PPB	5/23/2024 12:17	SMPL
P2550-13	MBH5X5	10654	2.2876 -		SFAM01.1	PPB	5/23/2024 12:19	SMPL
P2550-14	MBH5X5D	12691	2.7229 -		SFAM01.1	PPB	5/23/2024 12:21	SMPL
P2550-15	MBH5X5S	20436	4.3779 -		SFAM01.1	PPB	5/23/2024 12:23	SMPL
P2550-16	MBH5X6	5897	1.2711 -		SFAM01.1	PPB	5/23/2024 12:26	SMPL
P2550-17	MBH5X7	2175	0.4758 -		SFAM01.1	PPB	5/23/2024 12:28	SMPL
P2550-18	MBH5X8	9749	2.0942 -		SFAM01.1	PPB	5/23/2024 12:30	SMPL
P2550-19	MBH5X9	14204	3.0462 -		SFAM01.1	PPB	5/23/2024 12:33	SMPL
CCV034	CCV034	22775	4.8777 -		SFAM01.1	PPB	5/23/2024 12:35	SMPL
CCB034	CCB034	-160	-0.0232 -		SFAM01.1	PPB	5/23/2024 12:37	SMPL
P2550-20	MBH5Y0	15181	3.255 -		SFAM01.1	PPB	5/23/2024 12:39	SMPL
P2550-21	MBH5Y1	14853	3.1849 -		SFAM01.1	PPB	5/23/2024 12:42	SMPL
P2550-22	MBH5Y2	7585	1.6318 -		SFAM01.1	PPB	5/23/2024 12:44	SMPL
PB161091BL	PBS091	-52	-0.0001 -		SFAM01.1	PPB	5/23/2024 12:46	SMPL
P2551-01	MBH5Y3	8323	1.7895 -		SFAM01.1	PPB	5/23/2024 12:49	SMPL
P2551-02	MBH5Y4	11586	2.4868 -		SFAM01.1	PPB	5/23/2024 12:51	SMPL
P2551-03	MBH5Y5	7763	1.6698 -		SFAM01.1	PPB	5/23/2024 12:53	SMPL
P2551-04	MBH5Y6	3617	0.7839 -		SFAM01.1	PPB	5/23/2024 12:55	SMPL
P2551-05	MBH5Q0	17464	3.7428 -		SFAM01.1	PPB	5/23/2024 12:58	SMPL
P2551-06	MBH5Q1	13241	2.8404 -		SFAM01.1	PPB	5/23/2024 13:00	SMPL
P2551-07	MBH5Q2	9108	1.9573 -		SFAM01.1	PPB	5/23/2024 13:02	SMPL
P2551-08	MBH5Q3	4917	1.0617 -		SFAM01.1	PPB	5/23/2024 13:05	SMPL
P2551-09	MBH5Y7	5141	1.1096 -		SFAM01.1	PPB	5/23/2024 13:07	SMPL
P2551-10	MBH608	14648	3.1411 -		SFAM01.1	PPB	5/23/2024 13:09	SMPL

LB130896

INSTRUMENT ID ; CV1

P2551-11	MBH609	23687	5.0725 -	SFAM01.1	PPB	5/23/2024 13:11 SMPL
P2551-12	MBH610	23516	5.036 -	SFAM01.1	PPB	5/23/2024 13:14 SMPL
P2551-13	MBH610D	20405	4.3712 -	SFAM01.1	PPB	5/23/2024 13:16 SMPL
P2551-14	MBH610S	35929	7.6885 -	SFAM01.1	PPB	5/23/2024 13:18 SMPL
P2551-15	MBH611	12749	2.7353 -	SFAM01.1	PPB	5/23/2024 13:21 SMPL
P2551-16	MBH612	6076	1.3094 -	SFAM01.1	PPB	5/23/2024 13:23 SMPL
CCV035	CCV035	23306	4.9911 -	SFAM01.1	PPB	5/23/2024 13:25 SMPL
CCB035	CCB035	-138	-0.0185 -	SFAM01.1	PPB	5/23/2024 13:28 SMPL
P2551-17	MBH630	27216	5.8266 -	SFAM01.1	PPB	5/23/2024 13:30 SMPL
P2551-18	MBH5P9	10739	2.3058 -	SFAM01.1	PPB	5/23/2024 13:32 SMPL
P2551-19	MBH613	18951	4.0605 -	SFAM01.1	PPB	5/23/2024 13:34 SMPL
P2551-20	MBH614	18194	3.8988 -	SFAM01.1	PPB	5/23/2024 13:37 SMPL
P2551-21	MBH615	23815	5.0999 -	SFAM01.1	PPB	5/23/2024 13:39 SMPL
P2551-22	MBH616	14451	3.099 -	SFAM01.1	PPB	5/23/2024 13:41 SMPL
CCV036	CCV036	22739	4.87 -	SFAM01.1	PPB	5/23/2024 13:44 SMPL
CCB036	CCB036	-165	-0.0242 -	SFAM01.1	PPB	5/23/2024 13:46 SMPL