


LB130916

SFAM01.1 INSTRUMENT ID: CV1



Linear

A= 0.0000e+000
 B= 2.1591e-004 slope
 C= 2.0670e-002 y=mx+c
 Rho= 0.9998533
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	A/D
0.00	0.000	0.028	0.028	36	0.000	36					
0.05	0.050					442					
0.20	0.200	0.226	0.026	950	0.0 %	950	0				
2.50	2.500	2.515	0.015	11552	0.0 %	11552					
5.00	5.000	4.913	-0.087	22661	0.0 %	22661					
7.50	7.500	7.423	-0.077	34285	0.0 %	34285					
10.0	10.000	10.094	0.094	46656	0.0 %	46656					

A/D
 =
 13
 1
 -2
 -1
 1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	36	-		0 SFAM01.1	PPB	5/24/2024 10:07	Std
	0.2 S01	950	-		0.2 SFAM01.1	PPB	5/24/2024 10:09	Std
	2.5 S02	11552	-		2.5 SFAM01.1	PPB	5/24/2024 10:11	Std
	5 S03	22661	-		5 SFAM01.1	PPB	5/24/2024 10:14	Std
	7.5 S04	34285	-		7.5 SFAM01.1	PPB	5/24/2024 10:16	Std
	10 S05	46656	-		10 SFAM01.1	PPB	5/24/2024 10:18	Std
ICV062	ICV062	18433	4.0006 -		SFAM01.1	PPB	5/24/2024 10:21	SMPL
ICB062	ICB062	-134	-0.0083 -		SFAM01.1	PPB	5/24/2024 10:24	SMPL
CCV037	CCV037	23609	5.1181 -		SFAM01.1	PPB	5/24/2024 10:26	SMPL
CCB037	CCB037	-179	-0.018 -		SFAM01.1	PPB	5/24/2024 10:28	SMPL
PB161100BL	PBS100	153	0.0537 -		SFAM01.1	PPB	5/24/2024 10:30	SMPL
P2572-01	MBH5Q4	28449	6.1631 -		SFAM01.1	PPB	5/24/2024 10:33	SMPL
P2572-02	MBH5Q5	31377	6.7953 -		SFAM01.1	PPB	5/24/2024 10:35	SMPL
P2572-03	MBH5Q6	24761	5.3669 -		SFAM01.1	PPB	5/24/2024 10:37	SMPL
P2572-04	MBH5Q6D	29420	6.3728 -		SFAM01.1	PPB	5/24/2024 10:40	SMPL
P2572-05	MBH5Q6S	37923	8.2087 -		SFAM01.1	PPB	5/24/2024 10:42	SMPL
P2572-06	MBH5Q7	11745	2.5566 -		SFAM01.1	PPB	5/24/2024 10:44	SMPL
P2572-07	MBH5Q8	8910	1.9444 -		SFAM01.1	PPB	5/24/2024 10:46	SMPL
P2572-08	MBH632	17110	3.7149 -		SFAM01.1	PPB	5/24/2024 10:49	SMPL
P2572-09	MBH633	27362	5.9284 -		SFAM01.1	PPB	5/24/2024 10:51	SMPL
P2572-10	MBH634	13413	2.9167 -		SFAM01.1	PPB	5/24/2024 10:53	SMPL
P2572-11	MBH635	8126	1.7752 -		SFAM01.1	PPB	5/24/2024 10:56	SMPL
P2572-12	MBH636	5012	1.1028 -		SFAM01.1	PPB	5/24/2024 10:58	SMPL
P2572-13	MBH637	16798	3.6476 -		SFAM01.1	PPB	5/24/2024 11:00	SMPL
P2572-14	MBH638	19424	4.2145 -		SFAM01.1	PPB	5/24/2024 11:03	SMPL
P2572-15	MBH6B8	28087	6.085 -		SFAM01.1	PPB	5/24/2024 11:05	SMPL
P2572-16	MBH640	10108	2.2031 -		SFAM01.1	PPB	5/24/2024 11:07	SMPL
P2572-17	MBH641	5682	1.2475 -		SFAM01.1	PPB	5/24/2024 11:09	SMPL
P2572-18	MBH642	16322	3.5448 -		SFAM01.1	PPB	5/24/2024 11:12	SMPL
P2572-19	MBH643	24540	5.3191 -		SFAM01.1	PPB	5/24/2024 11:14	SMPL
CCV038	CCV038	22630	4.9068 -		SFAM01.1	PPB	5/24/2024 11:16	SMPL
CCB038	CCB038	-241	-0.0314 -		SFAM01.1	PPB	5/24/2024 11:19	SMPL
P2572-20	MBH644	20757	4.5023 -		SFAM01.1	PPB	5/24/2024 11:21	SMPL
P2572-21	MBH645	18425	3.9988 -		SFAM01.1	PPB	5/24/2024 11:23	SMPL
P2572-22	MBH646	6457	1.4148 -		SFAM01.1	PPB	5/24/2024 11:26	SMPL
PB161101BL	PBS101	-86	0.0021 -		SFAM01.1	PPB	5/24/2024 11:28	SMPL
P2573-01	MBH639	18704	4.0591 -		SFAM01.1	PPB	5/24/2024 11:30	SMPL
P2573-02	MBH639D	18340	3.9805 -		SFAM01.1	PPB	5/24/2024 11:32	SMPL
P2573-03	MBH639S	32372	7.0102 -		SFAM01.1	PPB	5/24/2024 11:35	SMPL
P2573-04	MBH647	17836	3.8717 -		SFAM01.1	PPB	5/24/2024 11:37	SMPL
P2573-05	MBH648	20238	4.3903 -		SFAM01.1	PPB	5/24/2024 11:39	SMPL
P2573-06	MBH649	8189	1.7888 -		SFAM01.1	PPB	5/24/2024 11:42	SMPL
P2573-07	MBH650	11370	2.4756 -		SFAM01.1	PPB	5/24/2024 11:44	SMPL
P2573-08	MBH651	2371	0.5326 -		SFAM01.1	PPB	5/24/2024 11:46	SMPL
P2573-09	MBH6B9	20014	4.3419 -		SFAM01.1	PPB	5/24/2024 11:48	SMPL
P2573-10	MBH5R4	44158	9.5549 -		SFAM01.1	PPB	5/24/2024 11:51	SMPL

LB130916

INSTRUMENT ID : CV1

P2573-11	MBH5R5	38341	8.2989 -	SFAM01.1	PPB	5/24/2024 11:53	SMPL
P2573-12	MBH5R7	21740	4.7146 -	SFAM01.1	PPB	5/24/2024 11:55	SMPL
P2573-13	MBH5R8	4658	1.0264 -	SFAM01.1	PPB	5/24/2024 11:58	SMPL
P2573-14	MBH697	46260	10.0087 -	SFAM01.1	PPB	5/24/2024 12:00	SMPL
P2573-15	MBH698	39833	8.6211 -	SFAM01.1	PPB	5/24/2024 12:02	SMPL
P2573-16	MBH699	26576	5.7587 -	SFAM01.1	PPB	5/24/2024 12:05	SMPL
CCV039	CCV039	22844	4.953 -	SFAM01.1	PPB	5/24/2024 12:07	SMPL
CCB039	CCB039	-221	-0.027 -	SFAM01.1	PPB	5/24/2024 12:09	SMPL
P2573-17	MBH6A0	23407	5.0745 -	SFAM01.1	PPB	5/24/2024 12:12	SMPL
P2573-18	MBH6A1	16957	3.6819 -	SFAM01.1	PPB	5/24/2024 12:14	SMPL
P2573-19	MBH6A2	30568	6.6207 -	SFAM01.1	PPB	5/24/2024 12:16	SMPL
P2573-20	MBH6A3	34467	7.4625 -	SFAM01.1	PPB	5/24/2024 12:18	SMPL
P2573-21	MBH6C0	81426	17.6015 -	SFAM01.1	PPB	5/24/2024 12:21	SMPL
P2573-22	MBH6C1	32240	6.9817 -	SFAM01.1	PPB	5/24/2024 12:23	SMPL
P2573-14DLX2	MBH697	22714	4.9249 -	SFAM01.1	PPB	5/24/2024 12:25	SMPL
P2573-21DLX5	MBH6C0	18708	4.0599 -	SFAM01.1	PPB	5/24/2024 12:28	SMPL
CCV040	CCV040	22518	4.8826 -	SFAM01.1	PPB	5/24/2024 12:30	SMPL
CCB040	CCB040	-276	-0.0389 -	SFAM01.1	PPB	5/24/2024 12:32	SMPL