

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/28/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:05
In Date: 05/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 05/25/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB130919

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P2590-01	BH6E7	1	1.14	8.69	9.83	8.1	80.1	
P2590-02	BHBK8	2	1.19	8.52	9.71	8.47	85.4	
P2590-03	BHBK9	3	1.19	8.63	9.82	8.71	87.1	
P2590-04	BHBL0	4	1.15	8.60	9.75	8.7	87.8	
P2590-05	BHBL1	5	1.13	8.68	9.81	8.65	86.6	
P2590-06	BHBL2	6	1.19	8.50	9.69	8.36	84.4	
P2590-07	BHBL3	7	1.15	8.40	9.55	8.12	83.0	
P2590-08	BHBL4	8	1.15	8.50	9.65	8.67	88.5	
P2590-09	BHBL5	9	1.13	8.73	9.86	8.34	82.6	
P2590-10	BHBL6	10	1.11	8.66	9.77	8.17	81.5	
P2590-11	BHBL7	11	1.19	8.43	9.62	8.03	81.1	
P2590-12	BHBL8	12	1.18	8.62	9.8	8.41	83.9	
P2590-13	BHBL9	13	1.16	8.61	9.77	8.66	87.1	
P2590-14	BHBM0	14	1.15	8.64	9.79	8.29	82.6	
P2590-15	BHBM0MS	15	1.15	8.64	9.79	8.29	82.6	
P2590-16	BHBM0MSD	16	1.15	8.64	9.79	8.29	82.6	
P2590-17	BHBM1	17	1.12	8.82	9.94	8.29	81.3	
P2590-18	BHBM2	18	1.19	8.63	9.82	8.4	83.5	
P2590-19	BHBM3	19	1.19	8.63	9.82	8.15	80.6	
P2590-20	BHBM4	20	1.15	8.35	9.5	8.5	88.0	
P2590-21	BHBM5	21	1.15	8.82	9.97	8.92	88.1	
P2590-22	BHBM6	22	1.19	8.73	9.92	8.91	88.4	
P2590-23	VHBLK001	23	1.00	1.00	2.00	2.00	100.0	vhblk

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$