


LB130926

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.1942e-004

C= 2.9671e-002

Rho= 0.9999525

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	7.11
0.0	0.000	0.036	0.036	27	0.000	27	0				
0.2	0.200	0.222	0.022	878	0.0 %	878					
2.5	2.500	2.426	-0.074	10919	0.0 %	10919					
5.0	5.000	5.001	0.001	22656	0.0 %	22656					
7.5	7.500	7.495	-0.005	34023	0.0 %	34023					
10.0	10.000	10.021	0.021	45534	0.0 %	45534					

7.11
11
3
0
0
0
0

LB130926 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 May 2024 13:48:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	51	27	-	0.0000	7470A	PPB	24 May 2024 13:51:46	S	Std
0.2 - 1	50.2	878	-	0.2000	7470A	PPB	24 May 2024 13:54:03	S	Std
2.5 - 1	52.5	10919	-	2.5000	7470A	PPB	24 May 2024 13:56:20	S	Std
5.0 - 1	55	22656	-	5.0000	7470A	PPB	24 May 2024 13:58:37	S	Std
7.5 - 1	59.5	34023	-	7.5000	7470A	PPB	24 May 2024 14:00:56	S	Std
10.0 - 1	110	45534	-	10.0000	7470A	PPB	24 May 2024 14:05:54	S	Std
ICV63 - 1	ICV63	17305	3.8267	-	7470A	PPB	24 May 2024 14:08:40	U	SMPL
ICB63 - 1	ICB63	-363	-0.0500	-	7470A	PPB	24 May 2024 14:10:56	U	SMPL
CCV41 - 1	CCV41	21929	4.8413	-	7470A	PPB	24 May 2024 14:13:14	U	SMPL
CCB41 - 1	CCB41	-463	-0.0719	-	7470A	PPB	24 May 2024 14:15:29	U	SMPL
CRA - 1	CRA	821	0.2098	-	7470A	PPB	24 May 2024 14:17:47	U	SMPL
HighStd - 1	HighStd	46359	10.2018	-	7470A	PPB	24 May 2024 14:20:02	U	SMPL
ChkStd - 1	ChkStd	29126	6.4205	-	7470A	PPB	24 May 2024 14:22:18	U	SMPL
PB161065BL - 1	PBW	-404	-0.0590	-	7470A	PPB	24 May 2024 14:24:41	U	SMPL
PB161065BS - 1	LCSW	15510	3.4329	-	7470A	PPB	24 May 2024 14:27:00	U	SMPL
P2556-02 - 1	COMP-1	-92	0.0095	-	7470A	PPB	24 May 2024 14:29:17	U	SMPL
P2558-02 - 1	COMP-1	244	0.0832	-	7470A	PPB	24 May 2024 14:31:36	U	SMPL
P2558-02DUP - 1	COMP-1DUP	282	0.0915	-	7470A	PPB	24 May 2024 14:33:54	U	SMPL
P2558-02MS - 1	COMP-1MS	15514	3.4337	-	7470A	PPB	24 May 2024 14:36:10	U	SMPL
P2558-02MSD - 1	COMP-1MSD	16324	3.6115	-	7470A	PPB	24 May 2024 14:38:26	U	SMPL
CCV42 - 1	CCV42	22364	4.9368	-	7470A	PPB	24 May 2024 14:40:43	U	SMPL
CCB42 - 1	CCB42	-537	-0.0882	-	7470A	PPB	24 May 2024 14:43:00	U	SMPL
PB161099BL - 1	PBW	-35	0.0220	-	7470A	PPB	24 May 2024 14:45:18	U	SMPL
PB161099BS - 1	LCSW	17886	3.9542	-	7470A	PPB	24 May 2024 14:47:33	U	SMPL
P2561-01 - 1	SY-2R	-268	-0.0291	-	7470A	PPB	24 May 2024 14:49:48	U	SMPL
P2561-02 - 1	SY-2R	-24	0.0244	-	7470A	PPB	24 May 2024 14:52:06	U	SMPL
P2561-03 - 1	SY-2D	-12	0.0270	-	7470A	PPB	24 May 2024 14:54:22	U	SMPL
P2561-04 - 1	SY-2D	30	0.0363	-	7470A	PPB	24 May 2024 14:56:38	U	SMPL
P2561-05 - 1	SY-7	141	0.0606	-	7470A	PPB	24 May 2024 14:58:54	U	SMPL
P2561-06 - 1	SY-7	31	0.0365	-	7470A	PPB	24 May 2024 15:01:12	U	SMPL
P2562-01 - 1	FIELD-BLANK	15	0.0330	-	7470A	PPB	24 May 2024 15:03:31	U	SMPL
P2562-02 - 1	FIELD-BLANK	5	0.0308	-	7470A	PPB	24 May 2024 15:05:47	U	SMPL
CCV43 - 1	CCV43	21948	4.8455	-	7470A	PPB	24 May 2024 15:08:03	U	SMPL
CCB43 - 1	CCB43	-263	-0.0280	-	7470A	PPB	24 May 2024 15:10:18	U	SMPL
P2562-03 - 1	SY-6	164	0.0657	-	7470A	PPB	24 May 2024 15:12:36	U	SMPL
P2562-04 - 1	SY-6	-49	0.0189	-	7470A	PPB	24 May 2024 15:14:52	U	SMPL
P2562-05 - 1	SY-3DD	42	0.0389	-	7470A	PPB	24 May 2024 15:17:07	U	SMPL
P2562-05DUP - 1	SY-3DDDUP	15	0.0330	-	7470A	PPB	24 May 2024 15:19:23	U	SMPL
P2562-06 - 1	P2562-05MS	16033	3.5476	-	7470A	PPB	24 May 2024 15:21:39	U	SMPL
P2562-07 - 1	P2562-05MSD	17681	3.9092	-	7470A	PPB	24 May 2024 15:23:55	U	SMPL
P2562-08 - 1	SY-3DD	-277	-0.0311	-	7470A	PPB	24 May 2024 15:26:14	U	SMPL
P2562-09 - 1	P2562-08MS	17636	3.8994	-	7470A	PPB	24 May 2024 15:28:34	U	SMPL
P2562-10 - 1	P2562-08MSD	15514	3.4337	-	7470A	PPB	24 May 2024 15:30:50	U	SMPL
P2562-11 - 1	SY-3D	1712	0.4053	-	7470A	PPB	24 May 2024 15:33:08	U	SMPL
CCV44 - 1	CCV44	24816	5.4748	-	7470A	PPB	24 May 2024 16:07:26	U	SMPL
CCB44 - 1	CCB44	-266	-0.0287	-	7470A	PPB	24 May 2024 16:11:53	U	SMPL
P2562-12 - 1	SY-3D	1391	0.3349	-	7470A	PPB	24 May 2024 16:14:11	U	SMPL
P2582-01 - 1	SY-3	2145	0.5003	-	7470A	PPB	24 May 2024 16:16:27	U	SMPL
P2582-02 - 1	SY-3	330	0.1021	-	7470A	PPB	24 May 2024 16:18:44	U	SMPL
PB161047TB - 1		116	0.0551	-	7470A	PPB	24 May 2024 16:21:02	U	SMPL
P2558-02LX5 - 1		4	0.0305	-	7470A	PPB	24 May 2024 16:23:18	U	SMPL
P2558-02A - 1		18597	4.1102	-	7470A	PPB	24 May 2024 16:25:35	U	SMPL
P2562-05LX5 - 1		-98	0.0082	-	7470A	PPB	24 May 2024 16:27:51	U	SMPL
P2562-05A - 1		18826	4.1605	-	7470A	PPB	24 May 2024 16:30:10	U	SMPL
CCV45 - 1	CCV45	23371	5.1577	-	7470A	PPB	24 May 2024 16:32:27	U	SMPL
CCB45 - 1	CCB45	-328	-0.0423	-	7470A	PPB	24 May 2024 16:34:45	U	SMPL