

66130929

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

5/24/2024 15:55

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.019	0.0	0.109	
ICB1	0.008	0.0	0.005	
CCV1	1.018	0.0	0.109	
CCB1	0.007	0.0	0.005	
PB161079BL	0.005	0.0	0.005	
PB161079BS	1.006	0.0	0.108	
P2403-01	0.032	0.0	0.007	
P2403-02	0.060	0.0	0.010	
PB161078BL	0.003	0.0	0.004	
PB161078BS	1.008	0.0	0.108	
P2403-07	0.027	0.0	0.007	
P2403-08	0.047	0.0	0.009	
P2561-01	0.008	0.0	0.005	
P2561-03	0.006	0.0	0.005	
CCV2	1.013	0.0	0.109	
CCB2	0.006	0.0	0.005	
P2561-05	0.006	0.0	0.005	
P2562-01	0.006	0.0	0.005	
P2562-03	0.007	0.0	0.005	
P2562-05	0.006	0.0	0.005	
P2562-05DUP	0.002	0.0	0.004	
P2562-06MS	1.038	0.0	0.111	
P2562-07MSD	1.050	0.0	0.113	
P2562-11	0.288	0.0	0.034	
P2582-01	0.437	0.0	0.049	
CCV3	1.018	0.0	0.109	
CCB3	0.007	0.0	0.005	

N	27
Mean	0.338
SD	0.4614
CV%	136.36

Aquakem v. 7.2AQ1

Results from time period:

Fri May 24 13:24:12 2024

Fri May 24 15:41:48 2024

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0036	mg/l	5/24/2024 14:01:32	
0.05PPM	A	Phenolics-	P	0.0568	mg/l	5/24/2024 14:01:33	
0.1PPM	A	Phenolics-	P	0.1037	mg/l	5/24/2024 14:01:34	
0.25PPM	A	Phenolics-	P	0.2509	mg/l	5/24/2024 14:01:35	
0.50PPM	A	Phenolics-	P	0.5001	mg/l	5/24/2024 14:01:36	
1.0PPM	A	Phenolics-	P	0.9851	mg/l	5/24/2024 14:01:37	
2.0PPM	A	Phenolics-	P	2.007	mg/l	5/24/2024 14:01:38	
ICV1	S	Phenolics-	P	1.0195	mg/l	5/24/2024 15:27:18	
ICB1	S	Phenolics-	P	0.0083	mg/l	5/24/2024 15:27:19	
CCV1	S	Phenolics-	P	1.0177	mg/l	5/24/2024 15:27:20	
CCB1	S	Phenolics-	P	0.0066	mg/l	5/24/2024 15:27:21	
PB161079BL	S	Phenolics-	P	0.0051	mg/l	5/24/2024 15:27:22	
PB161079BS	S	Phenolics-	P	1.0059	mg/l	5/24/2024 15:27:23	
P2403-01	S	Phenolics-	P	0.0317	mg/l	5/24/2024 15:27:24	
P2403-02	S	Phenolics-	P	0.0595	mg/l	5/24/2024 15:27:25	
PB161078BL	S	Phenolics-	P	0.003	mg/l	5/24/2024 15:27:26	
PB161078BS	S	Phenolics-	P	1.008	mg/l	5/24/2024 15:27:27	
P2403-07	S	Phenolics-	P	0.027	mg/l	5/24/2024 15:27:28	
P2403-08	S	Phenolics-	P	0.0472	mg/l	5/24/2024 15:27:29	
P2561-01	S	Phenolics-	P	0.0078	mg/l	5/24/2024 15:37:20	
P2561-03	S	Phenolics-	P	0.0057	mg/l	5/24/2024 15:37:21	
CCV2	S	Phenolics-	P	1.0131	mg/l	5/24/2024 15:37:22	
CCB2	S	Phenolics-	P	0.0056	mg/l	5/24/2024 15:37:23	
P2561-05	S	Phenolics-	P	0.0061	mg/l	5/24/2024 15:37:24	
P2562-01	S	Phenolics-	P	0.006	mg/l	5/24/2024 15:37:25	
P2562-03	S	Phenolics-	P	0.0071	mg/l	5/24/2024 15:37:26	
P2562-05	S	Phenolics-	P	0.0059	mg/l	5/24/2024 15:37:27	
P2562-05DUP	S	Phenolics-	P	0.0017	mg/l	5/24/2024 15:37:28	
P2562-06MS	S	Phenolics-	P	1.038	mg/l	5/24/2024 15:37:29	
P2562-07MSD	S	Phenolics-	P	1.0504	mg/l	5/24/2024 15:37:30	
P2562-11	S	Phenolics-	P	0.2879	mg/l	5/24/2024 15:37:31	
P2582-01	S	Phenolics-	P	0.4369	mg/l	5/24/2024 15:41:46	
CCV3	S	Phenolics-	P	1.0178	mg/l	5/24/2024 15:41:47	
CCB3	S	Phenolics-	P	0.0069	mg/l	5/24/2024 15:41:48	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

5/24/2024 14:02


Test Phenolics-

Accepted 5/24/2024 14:02

Factor 9.676
Bias 0.004

Coeff. of det. 0.999890

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.004	-0.0036	0.0000	-
2	Phenol-2	0.010	0.0568	0.0500	13.6
3	Phenol-2	0.015	0.1037	0.1000	3.7
4	Phenol-2	0.030	0.2509	0.2500	0.4
5	Phenol-2	0.056	0.5001	0.5000	0.0
6	Phenol-2	0.106	0.9851	1.0000	-1.5
7	Phenol-2	0.211	2.0070	2.0000	0.4

05/24/2024
RM