

LB130931

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/24/2024 16:31

Reviewed by : RM Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.012	0.0	0.109	
ICB1	0.005	0.0	0.005	
CCV1	1.005	0.0	0.108	
CCB1	0.001	0.0	0.004	
PB161087BL	0.003	0.0	0.004	
PB161087BS	1.008	0.0	0.108	
P2403-10	0.029	0.0	0.007	
P2403-11	0.052	0.0	0.009	
CCV2	1.011	0.0	0.109	
CCB2	0.002	0.0	0.004	
N	10			
Mean	0.413			
SD	0.5133			
CV%	124.35			

Aquakem v. 7.2AQ1

Results from time period:

Fri May 24 16:22:50 2024

Fri May 24 16:29:43 2024

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0036	mg/l	5/24/2024 14:01:32	
0.05PPM	A	Phenolics-	P	0.0568	mg/l	5/24/2024 14:01:33	
0.1PPM	A	Phenolics-	P	0.1037	mg/l	5/24/2024 14:01:34	
0.25PPM	A	Phenolics-	P	0.2509	mg/l	5/24/2024 14:01:35	
0.50PPM	A	Phenolics-	P	0.5001	mg/l	5/24/2024 14:01:36	
1.0PPM	A	Phenolics-	P	0.9851	mg/l	5/24/2024 14:01:37	
2.0PPM	A	Phenolics-	P	2.007	mg/l	5/24/2024 14:01:38	
ICV1	S	Phenolics-	P	1.0118	mg/l	5/24/2024 16:22:51	
ICB1	S	Phenolics-	P	0.0049	mg/l	5/24/2024 16:22:52	
CCV1	S	Phenolics-	P	1.0051	mg/l	5/24/2024 16:22:55	
CCB1	S	Phenolics-	P	0.0008	mg/l	5/24/2024 16:22:56	
PB161087BL	S	Phenolics-	P	0.0034	mg/l	5/24/2024 16:22:59	
PB161087BS	S	Phenolics-	P	1.0077	mg/l	5/24/2024 16:23:01	
P2403-10	S	Phenolics-	P	0.0291	mg/l	5/24/2024 16:29:37	
P2403-11	S	Phenolics-	P	0.052	mg/l	5/24/2024 16:29:39	
CCV2	S	Phenolics-	P	1.0106	mg/l	5/24/2024 16:29:40	
CCB2	S	Phenolics-	P	0.0019	mg/l	5/24/2024 16:29:43	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/24/2024 14:02

Reviewed by : RM Instrument ID : Konelab


Test Phenolics-

Accepted 5/24/2024 14:02

Factor 9.676
Bias 0.004

Coeff. of det. 0.999890

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.004	-0.0036	0.0000	-
2	Phenol-2	0.010	0.0568	0.0500	13.6
3	Phenol-2	0.015	0.1037	0.1000	3.7
4	Phenol-2	0.030	0.2509	0.2500	0.4
5	Phenol-2	0.056	0.5001	0.5000	0.0
6	Phenol-2	0.106	0.9851	1.0000	-1.5
7	Phenol-2	0.211	2.0070	2.0000	0.4

05/24/2024
RM