


LB130933

SFAM01.1 / INSTRUMENT ID: CV1

Linear

 μ Abs.:

49752



A= 0.0000e+000

B= 2.0440e-004 *slope*C= -8.9981e-002 *y-intercept*

Rho= 0.9998459

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.00	0.000	0.029	0.029	582	0.000	582					=
0.05	0.050					442	0				10
0.20	0.200	0.219	0.019	1512	0.0 %	1512					1
2.50	2.500	2.534	0.034	12840	0.0 %	12840					-2
5.00	5.000	4.878	-0.122	24304	0.0 %	24304					-1
7.50	7.500	7.461	-0.039	36941	0.0 %	36941					1
10.0	10.000	10.079	0.079	49752	0.0 %	49752					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	582	-		0 SFAM01.1	PPB	5/24/2024 17:30	Std
	0.2 S01	1512	-		0.2 SFAM01.1	PPB	5/24/2024 17:32	Std
	2.5 S02	12840	-		2.5 SFAM01.1	PPB	5/24/2024 17:34	Std
	5 S03	24304	-		5 SFAM01.1	PPB	5/24/2024 17:37	Std
	7.5 S04	36941	-		7.5 SFAM01.1	PPB	5/24/2024 17:39	Std
	10 S05	49752	-		10 SFAM01.1	PPB	5/24/2024 17:41	Std
ICV064	ICV064	19961	3.99 -		SFAM01.1	PPB	5/24/2024 17:44	SMPL
ICB064	ICB064	807	0.075 -		SFAM01.1	PPB	5/24/2024 17:46	SMPL
CCV046	CCV046	24356	4.8883 -		SFAM01.1	PPB	5/24/2024 17:49	SMPL
CCB046	CCB046	783	0.0701 -		SFAM01.1	PPB	5/24/2024 17:51	SMPL
PB161122BL	PBS122	325	-0.0236 -		SFAM01.1	PPB	5/24/2024 17:56	SMPL
P2574-01	MBH6A4	38658	7.8116 -		SFAM01.1	PPB	5/24/2024 17:58	SMPL
P2574-02	MBH6A4D	32893	6.6332 -		SFAM01.1	PPB	5/24/2024 18:00	SMPL
P2574-03	MBH6A4S	48130	9.7476 -		SFAM01.1	PPB	5/24/2024 18:03	SMPL
P2574-04	MBH6A5	28190	5.672 -		SFAM01.1	PPB	5/24/2024 18:05	SMPL
P2574-05	MBH6A6	30228	6.0885 -		SFAM01.1	PPB	5/24/2024 18:07	SMPL
P2574-06	MBH6A7	45529	9.216 -		SFAM01.1	PPB	5/24/2024 18:10	SMPL
P2574-07	MBH6A8	50545	10.2413 -		SFAM01.1	PPB	5/24/2024 18:12	SMPL
P2574-08	MBH6A9	35455	7.1569 -		SFAM01.1	PPB	5/24/2024 18:14	SMPL
P2574-09	MBH6B0	19458	3.8872 -		SFAM01.1	PPB	5/24/2024 18:17	SMPL
P2574-10	MBH6B1	10687	2.0944 -		SFAM01.1	PPB	5/24/2024 18:19	SMPL
P2574-11	MBH6B2	42065	8.508 -		SFAM01.1	PPB	5/24/2024 18:21	SMPL
P2574-12	MBH6B3	42639	8.6253 -		SFAM01.1	PPB	5/24/2024 18:23	SMPL
P2574-13	MBH6B4	78134	15.8804 -		SFAM01.1	PPB	5/24/2024 18:26	SMPL
P2574-14	MBH6B5	52276	10.5951 -		SFAM01.1	PPB	5/24/2024 18:28	SMPL
P2574-15	MBH6B6	52371	10.6145 -		SFAM01.1	PPB	5/24/2024 18:30	SMPL
P2574-16	MBH5Q9	25599	5.1424 -		SFAM01.1	PPB	5/24/2024 18:33	SMPL
P2574-17	MBH5R0	20209	4.0407 -		SFAM01.1	PPB	5/24/2024 18:35	SMPL
P2574-18	MBH5R2	18535	3.6985 -		SFAM01.1	PPB	5/24/2024 18:37	SMPL
P2574-19	MBH5R3	14522	2.8783 -		SFAM01.1	PPB	5/24/2024 18:40	SMPL
CCV047	CCV047	24131	4.8423 -		SFAM01.1	PPB	5/24/2024 18:42	SMPL
CCB047	CCB047	651	0.0431 -		SFAM01.1	PPB	5/24/2024 18:44	SMPL
P2574-20	MBH652	38578	7.7952 -		SFAM01.1	PPB	5/24/2024 18:47	SMPL
P2574-21	MBH653	29881	6.0176 -		SFAM01.1	PPB	5/24/2024 18:49	SMPL
P2574-22	MBH654	30560	6.1564 -		SFAM01.1	PPB	5/24/2024 18:51	SMPL
PB161123BL	PBS123	-212	-0.1333 -		SFAM01.1	PPB	5/24/2024 18:53	SMPL
P2575-01	MBH5R1	14871	2.9496 -		SFAM01.1	PPB	5/24/2024 18:56	SMPL
P2575-02	MBH5R1D	13889	2.7489 -		SFAM01.1	PPB	5/24/2024 18:58	SMPL
P2575-03	MBH5R1S	25918	5.2076 -		SFAM01.1	PPB	5/24/2024 19:00	SMPL
P2575-04	MBH655	23637	4.7413 -		SFAM01.1	PPB	5/24/2024 19:03	SMPL
P2575-05	MBH656	10076	1.9695 -		SFAM01.1	PPB	5/24/2024 19:05	SMPL
P2575-06	MBH657	34407	6.9427 -		SFAM01.1	PPB	5/24/2024 19:07	SMPL
P2575-07	MBH658	36239	7.3172 -		SFAM01.1	PPB	5/24/2024 19:10	SMPL
P2575-08	MBH659	46943	9.505 -		SFAM01.1	PPB	5/24/2024 19:12	SMPL
P2575-09	MBH660	24820	4.9831 -		SFAM01.1	PPB	5/24/2024 19:14	SMPL
P2575-10	MBH661	8813	1.7114 -		SFAM01.1	PPB	5/24/2024 19:17	SMPL

LB130933

INSTRUMENT ID : CV1

P2575-11	MBH662	26093	5.2433 -	SFAM01.1	PPB	5/24/2024 19:19 SMPL
P2575-12	MBH663	8041	1.5536 -	SFAM01.1	PPB	5/24/2024 19:21 SMPL
P2575-13	MBH664	25279	5.077 -	SFAM01.1	PPB	5/24/2024 19:23 SMPL
P2575-14	MBH665	26443	5.3149 -	SFAM01.1	PPB	5/24/2024 19:26 SMPL
P2575-15	MBH666	11564	2.2737 -	SFAM01.1	PPB	5/24/2024 19:28 SMPL
P2575-16	MBH667	33629	6.7837 -	SFAM01.1	PPB	5/24/2024 19:30 SMPL
CCV048	CCV048	24629	4.9441 -	SFAM01.1	PPB	5/24/2024 19:33 SMPL
CCB048	CCB048	500	0.0122 -	SFAM01.1	PPB	5/24/2024 19:35 SMPL
P2575-17	MBH668	29615	5.9632 -	SFAM01.1	PPB	5/24/2024 19:37 SMPL
P2575-18	MBH669	28016	5.6364 -	SFAM01.1	PPB	5/24/2024 19:39 SMPL
P2575-19	MBH670	19610	3.9182 -	SFAM01.1	PPB	5/24/2024 19:42 SMPL
P2575-20	MBH671	2212	0.3621 -	SFAM01.1	PPB	5/24/2024 19:44 SMPL
P2575-22	MBH6C2	23851	4.7851 -	SFAM01.1	PPB	5/24/2024 19:46 SMPL
P2574-07DLX2	MBH6A8	26089	5.2425 -	SFAM01.1	PPB	5/24/2024 19:49 SMPL
P2574-13DLX5	MBH6B5	16692	3.3218 -	SFAM01.1	PPB	5/24/2024 19:51 SMPL
P2574-14DLX2	MBH6B6	27211	5.4719 -	SFAM01.1	PPB	5/24/2024 19:53 SMPL
P2574-15DLX2	MBH6B4	27367	5.5037 -	SFAM01.1	PPB	5/24/2024 19:56 SMPL
CCV049	CCV049	25264	5.0739 -	SFAM01.1	PPB	5/24/2024 19:58 SMPL
CCB049	CCB049	529	0.0181 -	SFAM01.1	PPB	5/24/2024 20:00 SMPL