

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130938

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/28/2024	07:35
2	CAL2	1	Water	NA	NA	20.3	7.01	05/28/2024	07:36
3	CAL3	1	Water	NA	NA	20.3	10.03	05/28/2024	07:37
4	ICV	1	Water	NA	NA	20.3	7.00	05/28/2024	07:39
5	CCV1	1	Water	NA	NA	20.3	2.01	05/28/2024	07:40
6	P2611-01	1	Solid	20.02	20	21.7	7.84	05/28/2024	07:45
7	P2611-01DUP	1	Solid	20.03	20	21.9	7.85	05/28/2024	07:46
8	CCV2	1	Water	NA	NA	20.3	12.02	05/28/2024	07:50

WORKLIST(Hardcopy Internal Chain)

VB0936

WorkList Name : ph p2511 solids WorkList ID : 180559 Department : Wet-Chemistry Date : 05-28-2024 07:20:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2611-01	AU-06-052424	Solid	pH	Cool 4 deg C	PSEG05	K21	05/24/2024	9045D

Date/Time 05/28/24 07:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 05/28/24 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]