

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130939

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	05/28/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	05/28/2024	08:01
3	CAL3	1	Water	NA	NA	20.3	10.02	05/28/2024	08:05
4	ICV	1	Water	NA	NA	20.3	7.01	05/28/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.01	05/28/2024	08:12
6	P2607-02	1	Water	NA	NA	20.6	8.74	05/28/2024	08:25
7	P2607-02DUP	1	Water	NA	NA	20.7	8.75	05/28/2024	08:26
8	P2607-03	1	Water	NA	NA	20.7	4.64	05/28/2024	08:35
9	P2617-03	1	Water	NA	NA	23.2	5.25	05/28/2024	08:40
10	CCV2	1	Water	NA	NA	20.2	12.02	05/28/2024	08:41

WORKLIST(Hardcopy Internal Chain)

UB 130939

WorkList Name : ph p2617 water WorkList ID : 180560 Department : Wet-Chemistry Date : 05-28-2024 07:21:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2607-02	525	Water	pH	Cool 4 deg C	PSEG03	K21	05/24/2024	9040C
P2607-03	526	Water	pH	Cool 4 deg C	PSEG03	K21	05/24/2024	9040C
P2617-03	50360	Water	pH	Cool 4 deg C	PSEG03	K31	05/24/2024	9040C

Date/Time 05/28/24 07:30
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: JDCSM

Date/Time 05/28/24 11:00
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: JDCSM