

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130950

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/28/2024	14:25
2	CAL2	1	Water	NA	NA	20.2	7.00	05/28/2024	14:26
3	CAL3	1	Water	NA	NA	20.3	10.03	05/28/2024	14:30
4	ICV	1	Water	NA	NA	20.3	7.00	05/28/2024	14:33
5	CCV1	1	Water	NA	NA	20.4	2.01	05/28/2024	14:35
6	P2636-01	1	Water	NA	NA	22.0	8.15	05/28/2024	14:40
7	P2636-01DUP	1	Water	NA	NA	22.1	8.16	05/28/2024	14:41
8	CCV2	1	Water	NA	NA	20.5	12.03	05/28/2024	14:45

WORKLIST(Hardcopy Internal Chain)

130950

WorkList Name : PH P2636

WorkList ID : 180592

Department : Wet-Chemistry

Date : 05-28-2024 14:11:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2636-01	TOTE-COMP	Water	pH	Cool 4 deg C	PSEG03	K21	05/28/2024	9040C

Date/Time 05/28/24 14:20
Raw Sample Received by: J. C. (sm)
Raw Sample Relinquished by: J. C. (sm)

Date/Time 05/28/24 16:30
Raw Sample Received by: J. C. (sm)
Raw Sample Relinquished by: J. C. (sm)