

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130957

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/28/2024	15:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/28/2024	15:16
3	CAL3	1	Water	NA	NA	20.3	10.03	05/28/2024	15:17
4	ICV	1	Water	NA	NA	20.4	7.02	05/28/2024	15:19
5	CCV1	1	Water	NA	NA	20.3	2.01	05/28/2024	15:22
6	P2612-03	1	Solid	20.02	20	22.0	8.36	05/28/2024	15:30
7	P2612-06	1	Solid	20.03	20	22.1	7.29	05/28/2024	15:33
8	P2612-09	1	Solid	20.04	20	21.9	8.66	05/28/2024	15:37
9	P2612-12	1	Solid	20.02	20	22.2	7.02	05/28/2024	15:45
10	P2637-04	1	Solid	20.03	20	21.4	8.35	05/28/2024	15:50
11	P2637-04DUP	1	Solid	20.04	20	21.5	8.36	05/28/2024	15:51
12	CCV2	1	Water	NA	NA	20.3	12.03	05/28/2024	16:00

WORKLIST(Hardcopy Internal Chain)

130957

WorkList Name : corrosivity p2612 WorkList ID : 180572 Department : Wet-Chemistry Date : 05-28-2024 10:09:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2612-03	VNJ-203	Solid	Corrosivity	Cool 4 deg C	PSEG03	K21	05/24/2024	9045D
P2612-06	RT1486	Solid	Corrosivity	Cool 4 deg C	PSEG03	K21	05/24/2024	9045D
P2612-09	RBR200057	Solid	Corrosivity	Cool 4 deg C	PSEG03	K21	05/24/2024	9045D
P2612-12	72-11933	Solid	Corrosivity	Cool 4 deg C	PSEG03	K21	05/24/2024	9045D
P2637-04	WCS-TPT	Solid	Corrosivity	Cool 4 deg C	PSEG03	K21	05/28/2024	9045D

Date/Time 05/28/24 15:00
 Raw Sample Received by: 20620
 Raw Sample Relinquished by: JDCSM

Date/Time 05/28/24 17:00
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM