

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB130962

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP108228
Turbidity Calibration std, 0NTU	WP108220
Turbidity Calibration std, 20NTU	WP108221
Turbidity Calibration std, 40NTU	WP108222
Turbidity Calibration std, 1NTU	WP108224
Turbidity Calibration std, 5NTU	WP108225
Turbidity Calibration - CCV std, 10 NTU	WP108227
10 NTU Standard 500 ml	W2971

Intercept: -0.0556

Slope: 0.9578

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.21		05/29/2024	13:30
2	CAL2	1	1	1.120	1.23	22.7	05/29/2024	13:34
3	CAL3	5	1	4.400	4.65	-7	05/29/2024	13:38
4	CAL4	10	1	9.490	9.97	-0.3	05/29/2024	13:42
5	CAL5	20	1	18.900	19.79	-1	05/29/2024	13:46
6	CAL6	40	1	38.400	40.15	0.4	05/29/2024	13:50



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB130962

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.500	9.977	05/29/2024	13:54
2	ICB		1	0.150	0.215	05/29/2024	13:58
3	CCV1	10	1	9.500	9.977	05/29/2024	14:02
4	CCB1		1	0.150	0.215	05/29/2024	14:06
5	LB130962BL		1	0.140	0.204	05/29/2024	14:10
6	P2403-09		1	0.510	0.591	05/29/2024	14:14
7	CCV2	10	1	9.500	9.977	05/29/2024	14:18
8	CCB2		1	0.150	0.215	05/29/2024	14:22