


LB 130963

SFAM01.1

INSTRUMENT ID: CY1



Linear

A= 0.0000e+000

B= 2.3650e-004 *slope*C= 3.1282e-002 *y-intercept*

Rho= 0.9997013

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>G/P</i>
0.00	0.000	0.044	0.044	54	0.000	54					0
0.05	0.050					442	0				1
0.20	0.200	0.214	0.014	771	0.0 %	771					2
2.50	2.500	2.335	-0.165	9742	0.0 %	9742					2
5.00	5.000	5.123	0.123	21530	0.0 %	21530					1
7.50	7.500	7.538	0.038	31742	0.0 %	31742					1
10.0	10.000	9.946	-0.054	41923	0.0 %	41923					-1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	54	-		0 SFAM01.1	PPB	5/29/2024 14:06	Std
	0.2 S01	771	-		0.2 SFAM01.1	PPB	5/29/2024 14:09	Std
	2.5 S02	9742	-		2.5 SFAM01.1	PPB	5/29/2024 14:11	Std
	5 S03	21530	-		5 SFAM01.1	PPB	5/29/2024 14:13	Std
	7.5 S04	31742	-		7.5 SFAM01.1	PPB	5/29/2024 14:17	Std
	10 S05	41923	-		10 SFAM01.1	PPB	5/29/2024 14:22	Std
ICV066	ICV066	16752	3.9931 -		SFAM01.1	PPB	5/29/2024 14:25	SMPL
ICB066	ICB066	-151	-0.0044 -		SFAM01.1	PPB	5/29/2024 14:27	SMPL
CCV054	CCV054	21685	5.1597 -		SFAM01.1	PPB	5/29/2024 14:30	SMPL
CCB054	CCB054	-192	-0.0141 -		SFAM01.1	PPB	5/29/2024 14:32	SMPL
PB161183BL	PBS183	16	0.0351 -		SFAM01.1	PPB	5/29/2024 14:34	SMPL
P2552-01	MBH617	8508	2.0434 -		SFAM01.1	PPB	5/29/2024 14:36	SMPL
P2552-02	MBH618	18238	4.3445 -		SFAM01.1	PPB	5/29/2024 14:39	SMPL
P2552-03	MBH619	17550	4.1818 -		SFAM01.1	PPB	5/29/2024 14:41	SMPL
P2552-04	MBH620	19784	4.7101 -		SFAM01.1	PPB	5/29/2024 14:43	SMPL
P2552-05	MBH621	20004	4.7622 -		SFAM01.1	PPB	5/29/2024 14:45	SMPL
P2552-06	MBH622	6831	1.6468 -		SFAM01.1	PPB	5/29/2024 14:48	SMPL
P2552-07	MBH623	26458	6.2885 -		SFAM01.1	PPB	5/29/2024 14:50	SMPL
P2552-08	MBH624	29456	6.9975 -		SFAM01.1	PPB	5/29/2024 14:52	SMPL
P2552-10	MBH625	24089	5.7282 -		SFAM01.1	PPB	5/29/2024 14:55	SMPL
P2552-11	MBH625D	25414	6.0416 -		SFAM01.1	PPB	5/29/2024 14:57	SMPL
P2552-12	MBH625S	37122	8.8105 -		SFAM01.1	PPB	5/29/2024 15:02	SMPL
P2552-13	MBH626	24514	5.8287 -		SFAM01.1	PPB	5/29/2024 15:04	SMPL
P2552-14	MBH627	20024	4.7669 -		SFAM01.1	PPB	5/29/2024 15:07	SMPL
P2552-15	MBH629	15824	3.7736 -		SFAM01.1	PPB	5/29/2024 15:09	SMPL
P2552-16	MBH631	28153	6.6894 -		SFAM01.1	PPB	5/29/2024 15:11	SMPL
P2552-17	MBHBM7	1655	0.4227 -		SFAM01.1	PPB	5/29/2024 15:13	SMPL
P2552-18	MBHBM8	9728	2.3319 -		SFAM01.1	PPB	5/29/2024 15:16	SMPL
P2552-19	MBHBM9	19431	4.6266 -		SFAM01.1	PPB	5/29/2024 15:18	SMPL
P2552-20	MBHBN0	11181	2.6755 -		SFAM01.1	PPB	5/29/2024 15:20	SMPL
CCV055	CCV055	23406	5.5667 -		SFAM01.1	PPB	5/29/2024 15:23	SMPL
CCB055	CCB055	-137	-0.0011 -		SFAM01.1	PPB	5/29/2024 15:25	SMPL
P2552-21	MBHBN1	55147	13.0733 -		SFAM01.1	PPB	5/29/2024 15:27	SMPL
P2552-22	MBHBN2	34666	8.2297 -		SFAM01.1	PPB	5/29/2024 15:29	SMPL
PB161184BL	PBS184	-220	-0.0207 -		SFAM01.1	PPB	5/29/2024 15:32	SMPL
P2576-04	MBHBN3	6555	1.5815 -		SFAM01.1	PPB	5/29/2024 15:41	SMPL
P2576-05	MBHBN4	3584	0.8789 -		SFAM01.1	PPB	5/29/2024 15:43	SMPL
P2576-06	MBHBN5	2266	0.5672 -		SFAM01.1	PPB	5/29/2024 15:46	SMPL
P2576-07	MBHBN6	15410	3.6757 -		SFAM01.1	PPB	5/29/2024 15:48	SMPL
P2576-08	MBHBN7	68094	16.1353 -		SFAM01.1	PPB	5/29/2024 15:50	SMPL
P2576-09	MBHBN8	4126	1.0071 -		SFAM01.1	PPB	5/29/2024 15:52	SMPL
P2576-10	MBHBN9	8973	2.1534 -		SFAM01.1	PPB	5/29/2024 15:55	SMPL
P2576-11	MBHBP0	25146	5.9782 -		SFAM01.1	PPB	5/29/2024 15:57	SMPL
P2576-12	MBHBP1	17505	4.1711 -		SFAM01.1	PPB	5/29/2024 15:59	SMPL
P2576-13	MBHBP2	6200	1.4976 -		SFAM01.1	PPB	5/29/2024 16:02	SMPL
P2576-14	MBHBP8	6563	1.5834 -		SFAM01.1	PPB	5/29/2024 16:04	SMPL

LB130963

INSTRUMENT ID : CV1

P2576-15	MBHBP9	25006	5.9451 -	SFAM01.1	PPB	5/29/2024 16:06 SMPL
P2576-16	MBHBQ0	4040	0.9867 -	SFAM01.1	PPB	5/29/2024 16:08 SMPL
P2576-17	MBHBW4	6455	1.5579 -	SFAM01.1	PPB	5/29/2024 16:11 SMPL
CCV056	CCV056	23370	5.5582 -	SFAM01.1	PPB	5/29/2024 16:13 SMPL
CCB056	CCB056	-107	0.006 -	SFAM01.1	PPB	5/29/2024 16:15 SMPL
P2576-18	MBHBR5	15348	3.661 -	SFAM01.1	PPB	5/29/2024 16:18 SMPL
P2576-19	MBHBR6	44460	10.5459 -	SFAM01.1	PPB	5/29/2024 16:20 SMPL
P2576-20	MBHBR7	122323	28.9602 -	SFAM01.1	PPB	5/29/2024 16:22 SMPL
P2576-21	MBHBR9	14013	3.3453 -	SFAM01.1	PPB	5/29/2024 16:24 SMPL
P2576-22	MBHBS0	15584	3.7168 -	SFAM01.1	PPB	5/29/2024 16:27 SMPL
P2576-01	MBH5R6	45611	10.8181 -	SFAM01.1	PPB	5/29/2024 16:29 SMPL
P2576-02	MBH5R6D	46703	11.0764 -	SFAM01.1	PPB	5/29/2024 16:35 SMPL
P2576-03	MBH5R6S	60415	14.3192 -	SFAM01.1	PPB	5/29/2024 16:42 SMPL
P2552-21DLX5	MBHBN1	10799	2.5852 -	SFAM01.1	PPB	5/29/2024 16:45 SMPL
P2576-08DLX5	MBHBN7	15999	3.815 -	SFAM01.1	PPB	5/29/2024 16:47 SMPL
P2576-01DLX2	MBH5R6	22976	5.465 -	SFAM01.1	PPB	5/29/2024 16:49 SMPL
P2576-02DLX2	MBH5R6D	23429	5.5722 -	SFAM01.1	PPB	5/29/2024 16:51 SMPL
P2576-03DLX2	MBH5R6S	32237	7.6552 -	SFAM01.1	PPB	5/29/2024 17:01 SMPL
P2576-19DLX2	MBHBR6	23421	5.5703 -	SFAM01.1	PPB	5/29/2024 17:04 SMPL
P2576-20DLX5	MBHBR7	25978	6.175 -	SFAM01.1	PPB	5/29/2024 17:09 SMPL
CCV057	CCV057	23940	5.693 -	SFAM01.1	PPB	5/29/2024 17:11 SMPL
CCB057	CCB057	-130	0.0005 -	SFAM01.1	PPB	5/29/2024 17:13 SMPL