

TitriSoft Report

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	slope	offset
LB130966BLW	100	30/05/2024 09:35	0.00	3.74	3.76	103.6	7.02
LB130966BSW	100	30/05/2024 09:38	52.36	4.38	8.64	103.6	7.02
P2561-01	50	30/05/2024 09:45	18.22	4.42	5.79	103.6	7.02
P2561-03	50	30/05/2024 09:49	68.56	4.38	6.18	103.6	7.02
P2561-05	50	30/05/2024 09:53	67.08	4.36	6.46	103.6	7.02
P2562-01	50	30/05/2024 09:58	2.28	4.26	5.41	103.6	7.02
P2562-03	50	30/05/2024 10:03	105.92	4.40	6.85	103.6	7.02
P2562-05	50	30/05/2024 10:07	4.74	4.42	6.46	103.6	7.02
P2562-11	50	30/05/2024 10:12	193.68	4.36	6.74	103.6	7.02
P2582-01	50	30/05/2024 10:18	104.88	4.41	6.68	103.6	7.02
P2582-01DUP	50	30/05/2024 10:22	103.04	4.42	6.67	103.6	7.02

05/30/2024

12
05/30/24

WORKLIST(Hardcopy Internal Chain)

LB130966

WorkList Name : ALKALINITY-5-24

WorkList ID : 180520

Department : Wet-Chemistry

Date : 05-24-2024 08:47:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2561-01	SY-2R	Water	Alkalinity	Cool 4 deg C	LOCK01	K31	05/21/2024	SM2320 B
P2561-03	SY-2D	Water	Alkalinity	Cool 4 deg C	LOCK01	K31	05/21/2024	SM2320 B
P2561-05	SY-7	Water	Alkalinity	Cool 4 deg C	LOCK01	K31	05/21/2024	SM2320 B
P2562-01	FIELD-BLANK	Water	Alkalinity	Cool 4 deg C	LOCK01	K21	05/20/2024	SM2320 B
P2562-03	SY-6	Water	Alkalinity	Cool 4 deg C	LOCK01	K21	05/20/2024	SM2320 B
P2562-05	SY-3DD	Water	Alkalinity	Cool 4 deg C	LOCK01	K21	05/20/2024	SM2320 B
P2562-11	SY-3D	Water	Alkalinity	Cool 4 deg C	LOCK01	K21	05/20/2024	SM2320 B
P2582-01	SY-3	Water	Alkalinity	Cool 4 deg C	LOCK01	K21	05/22/2024	SM2320 B

Date/Time

05/30/24 08:05

Raw Sample Received by:

12(1C)

Raw Sample Relinquished by:

4800(1)

Date/Time

05/30/24 10:40

Raw Sample Received by:

4800(1)

Raw Sample Relinquished by:

12(1C)