

LB130988


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46147



A= 0.0000e+000

B= 2.1708e-004 slope

C= -6.0523e-002 intercept

Rho= 0.9999055

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/v)
0.00	0.000	0.048	0.048	499	0.000	499					1
0.05	0.050					442	0				1
0.20	0.200	0.170	-0.030	1064	0.0 %	1064					1
2.50	2.500	2.472	-0.028	11666	0.0 %	11666					1
5.00	5.000	4.962	-0.038	23138	0.0 %	23138					1
7.50	7.500	7.591	0.091	35246	0.0 %	35246					1
10.0	10.000	9.957	-0.043	46147	0.0 %	46147					0

LB130988

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	499	-		0 SFAM01.1	PPB	5/30/2024 14:25	Std
	0.2 S01	1064	-		0.2 SFAM01.1	PPB	5/30/2024 14:27	Std
	2.5 S02	11666	-		2.5 SFAM01.1	PPB	5/30/2024 14:30	Std
	5 S03	23138	-		5 SFAM01.1	PPB	5/30/2024 14:32	Std
	7.5 S04	35246	-		7.5 SFAM01.1	PPB	5/30/2024 14:34	Std
	10 S05	46147	-		10 SFAM01.1	PPB	5/30/2024 14:37	Std
ICV068	ICV068	18155	3.8805 -		SFAM01.1	PPB	5/30/2024 14:40	SMPL
ICB068	ICB068	500	0.048 -		SFAM01.1	PPB	5/30/2024 14:42	SMPL
CCV061	CCV061	22316	4.7838 -		SFAM01.1	PPB	5/30/2024 14:44	SMPL
CCB061	CCB061	480	0.0437 -		SFAM01.1	PPB	5/30/2024 14:46	SMPL
PB161216BL	PBW216	-34	-0.0679 -		SFAM01.1	PPB	5/30/2024 14:49	SMPL
P2552-09	MBH5N8	-9	-0.0625 -		SFAM01.1	PPB	5/30/2024 14:51	SMPL
P2565-04	MX004	6892	1.4356 -		SFAM01.1	PPB	5/30/2024 14:55	SMPL
P2575-21	MBH5R9	-29	-0.0668 -		SFAM01.1	PPB	5/30/2024 15:00	SMPL
P2592-01	MBH5S0	-3	-0.0612 -		SFAM01.1	PPB	5/30/2024 15:02	SMPL
CCV062	CCV062	21722	4.6549 -		SFAM01.1	PPB	5/30/2024 15:04	SMPL
CCB062	CCB062	485	0.0448 -		SFAM01.1	PPB	5/30/2024 15:07	SMPL