

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB130992

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP108275
Residual chlorine ICV-LCS, 0.4PPM	WP108274
Chlorine Calibration std, 0.1ppm	WP108269
Chlorine Calibration std, 0.2ppm	WP108270
Chlorine Calibration std, 0.8ppm	WP108271
Chlorine Calibration std, 0.0ppm	WP108266
Chlorine Calibration std, 1.6ppm	WP108272
Residual Chlorine Calibration and CCV std, 0	WP108273
Total Chlorine Powder Pillows	W2815

Intercept: 0.0134

Slope: 0.9353

Regression: 0.999801

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		05/30/2024	14:00
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	3	05/30/2024	14:03
3	CAL3	0.2	1	0.000	0.200	0.200	0.20	0	05/30/2024	14:06
4	CAL4	0.4	1	0.000	0.390	0.390	0.40	0.8	05/30/2024	14:09
5	CAL5	0.8	1	0.000	0.780	0.780	0.82	2.5	05/30/2024	14:12
6	CAL6	1.6	1	0.000	1.500	1.500	1.59	-0.7	05/30/2024	14:15

Analysis Method: SM4500 Cl GANALYST: NihaParameter: Residual ChlorineSUPERVISOR REVIEW BY: IwonaRun Number: LB130992

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4000	0.4000	0.4130	05/30/2024	14:18
2	ICB				1	0.0000	0.0000	0.0000	-0.0140	05/30/2024	14:21
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4240	05/30/2024	14:24
4	CCB1				1	0.0000	0.0100	0.0100	-0.0040	05/30/2024	14:27
5	LB130992BL	50	50		1	0.0000	0.0000	0.0000	-0.0140	05/30/2024	14:30
6	LB130992BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.4130	05/30/2024	14:33
7	P2403-09	50	50		1	0.0000	0.0560	0.0560	0.0460	05/30/2024	14:36
8	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4240	05/30/2024	14:39
9	CCB2				1	0.0000	0.0000	0.0000	-0.0140	05/30/2024	14:42