

LB131015


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44282



A= 0.0000e+000

B= 2.2794e-004

C= 1.0950e-002

Rho= 0.9995933

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.086	0.086	331	0.000	331	0			
0.2	0.200	0.184	-0.016	760	0.0 %	760				
2.5	2.500	2.414	-0.086	10542	0.0 %	10542				
5.0	5.000	5.090	0.090	22284	0.0 %	22284				
7.5	7.500	7.321	-0.179	32068	0.0 %	32068				
10.0	10.000	10.105	0.105	44282	0.0 %	44282				

0.000
1
18
3
2
2
1

LB131015 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 May 2024 13:54:18

Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	331	-	0.0000	7470A	PPB	31 May 2024 13:56:23	S	Std
0.2 - 1	50.2	760	-	0.2000	7470A	PPB	31 May 2024 13:58:40	S	Std
2.5 - 1	52.5	10542	-	2.5000	7470A	PPB	31 May 2024 14:00:58	S	Std
5.0 - 1	55	22284	-	5.0000	7470A	PPB	31 May 2024 14:03:16	S	Std
7.5 - 1	57.5	32068	-	7.5000	7470A	PPB	31 May 2024 14:05:33	S	Std
10.0 - 1	510	44282	-	10.0000	7470A	PPB	31 May 2024 14:16:09	S	Std
ICV70 - 1	ICV70	18106	4.1380	-	7470A	PPB	31 May 2024 14:24:48	U	SMPL
ICB70 - 1	ICB70	43	0.0208	-	7470A	PPB	31 May 2024 14:27:03	U	SMPL
CCV66 - 1	CCV66	22176	5.0658	-	7470A	PPB	31 May 2024 14:29:19	U	SMPL
CCB66 - 1	CCB66	12	0.0137	-	7470A	PPB	31 May 2024 14:31:34	U	SMPL
CRA - 1	CRA	640	0.1568	-	7470A	PPB	31 May 2024 14:33:52	U	SMPL
HighStd - 1	HighStd	43349	9.8919	-	7470A	PPB	31 May 2024 14:36:08	U	SMPL
ChkStd - 1	ChkStd	28637	6.5385	-	7470A	PPB	31 May 2024 14:38:23	U	SMPL
PB161241BL - 1	PBW	-259	-0.0481	-	7470A	PPB	31 May 2024 14:40:42	U	SMPL
PB161241BS - 1	LCSW	15901	3.6354	-	7470A	PPB	31 May 2024 14:43:00	U	SMPL
P2607-02 - 1	525	101	0.0340	-	7470A	PPB	31 May 2024 14:45:16	U	SMPL
P2636-01 - 1	TOTE-COMP	121	0.0385	-	7470A	PPB	31 May 2024 14:48:33	U	SMPL
P2636-01DUP - 1	TOTE-COMPDUP	175	0.0508	-	7470A	PPB	31 May 2024 14:50:49	U	SMPL
P2636-01MS - 1	TOTE-COMPMS	15669	3.5825	-	7470A	PPB	31 May 2024 14:53:06	U	SMPL
P2636-01MSD - 1	TOTE-COMPMSD	15088	3.4501	-	7470A	PPB	31 May 2024 14:55:21	U	SMPL
CCV67 - 1	CCV67	22238	5.0799	-	7470A	PPB	31 May 2024 14:57:38	U	SMPL
CCB67 - 1	CCB67	82	0.0296	-	7470A	PPB	31 May 2024 14:59:55	U	SMPL
P2636-01LX5 - 1		103	0.0344	-	7470A	PPB	31 May 2024 15:02:13	U	SMPL
P2636-01A - 1		16458	3.7624	-	7470A	PPB	31 May 2024 15:04:28	U	SMPL
CCV68 - 1	CCV68	23153	5.2885	-	7470A	PPB	31 May 2024 15:06:45	U	SMPL
CCB68 - 1	CCB68	39	0.0198	-	7470A	PPB	31 May 2024 15:09:03	U	SMPL