

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131031

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	06/03/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.01	06/03/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.03	06/03/2024	08:14
4	ICV	1	Water	NA	NA	20.3	6.99	06/03/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	06/03/2024	08:20
6	P2691-02	1	Water	NA	NA	23.4	7.51	06/03/2024	08:25
7	P2691-02DUP	1	Water	NA	NA	23.5	7.52	06/03/2024	08:26
8	P2691-03	1	Water	NA	NA	23.3	10.73	06/03/2024	08:33
9	P2691-04	1	Water	NA	NA	23.4	5.52	06/03/2024	08:37
10	CCV2	1	Water	NA	NA	20.4	12.02	06/03/2024	08:40

WORKLIST(Hardcopy Internal Chain)

131031

WorkList Name : ph p2691 WorkList ID : 180729 Department : Wet-Chemistry Date : 06-03-2024 07:47:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2691-02	C RINSATE-1	Water	pH	Cool 4 deg C	PSEG03	K21	05/31/2024	9040C
P2691-03	C RINSATE-2	Water	pH	Cool 4 deg C	PSEG03	K21	05/31/2024	9040C
P2691-04	C 50362	Water	pH	Cool 4 deg C	PSEG03	K21	05/31/2024	9040C

Date/Time 06/03/24 08:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06/03/24 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]