

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131036

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/03/2024	11:30
2	CAL2	1	Water	NA	NA	20.3	7.01	06/03/2024	11:31
3	CAL3	1	Water	NA	NA	20.3	10.03	06/03/2024	11:35
4	ICV	1	Water	NA	NA	20.4	7.02	06/03/2024	11:37
5	CCV1	1	Water	NA	NA	20.3	2.01	06/03/2024	11:40
6	P2689-01	1	Solid	20.02	20	21.9	7.57	06/03/2024	11:47
7	P2689-01DUP	1	Solid	20.03	20	22.1	7.58	06/03/2024	11:48
8	P2691-01	1	Solid	20.01	20	21.3	5.09	06/03/2024	11:55
9	CCV2	1	Water	NA	NA	20.4	12.02	06/03/2024	12:00

WORKLIST(Hardcopy Internal Chain)

CP 131036

WorkList Name : ph p2689 WorkList ID : 180744 Department : Wet-Chemistry Date : 06-03-2024 11:09:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P2689-01	BU-03-05312024	Solid	pH	Cool 4 deg C	PSEG05	K21	05/31/2024	9045D
P2691-01	DIESEL-1	Solid	pH	Cool 4 deg C	PSEG03	K21	05/31/2024	9045D

Date/Time 06/03/24 11:20

Raw Sample Received by: rd 600

Raw Sample Relinquished by: CP 5

Date/Time 06/03/24 13:30

Raw Sample Received by: CP 5

Raw Sample Relinquished by: rd 600