


LB131040

7470A

INSTRUMENT ID: CV2

 μ Abs.:

41529



Linear

A= 0.0000e+000

B= 2.4228e-004

C= 4.9796e-002

Rho= 0.9995391

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	0.068	0.068	74	0.000	74	0				-
0.2	0.200	0.198	-0.002	613	0.0 %	613					-1
2.5	2.500	2.419	-0.081	9779	0.0 %	9779					-3
5.0	5.000	5.105	0.105	20866	0.0 %	20866					-2
7.5	7.500	7.298	-0.202	29918	0.0 %	29918					-3
10.0	10.000	10.111	0.111	41529	0.0 %	41529					-1

LB131040 IONSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jun 2024 15:41:10

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.000	7470A	PPB	03 Jun 2024 15:46:46	S	Std
0.2 - 1	50.2	613	-	0.200	7470A	PPB	03 Jun 2024 15:51:13	S	Std
2.5 - 1	52.5	9779	-	2.500	7470A	PPB	03 Jun 2024 15:53:30	S	Std
5.0 - 1	35	20866	-	5.000	7470A	PPB	03 Jun 2024 15:55:47	S	Std
7.5 - 1	37.5	29918	-	7.500	7470A	PPB	03 Jun 2024 15:58:03	S	Std
10.0 - 1	510	41529	-	10.000	7470A	PPB	03 Jun 2024 16:02:59	S	Std
ICV72 - 1	ICV72	15288	3.7538	-	7470A	PPB	03 Jun 2024 16:08:27	U	SMPL
ICB72 - 1	ICB72	-15	0.0462	-	7470A	PPB	03 Jun 2024 16:13:20	U	SMPL
CCV73 - 1	CCV73	21517	5.2629	-	7470A	PPB	03 Jun 2024 16:15:34	U	SMPL
CCB73 - 1	CCB73	-205	0.0001	-	7470A	PPB	03 Jun 2024 16:17:50	U	SMPL
CRA - 1	CRA	494	0.1695	-	7470A	PPB	03 Jun 2024 16:20:07	U	SMPL
HighStd - 1	HighStd	42857	10.4332	-	7470A	PPB	03 Jun 2024 16:22:22	U	SMPL
ChkStd - 1	ChkStd	28456	6.9441	-	7470A	PPB	03 Jun 2024 16:24:37	U	SMPL
PB161243BL - 1	PBW	163	0.0893	-	7470A	PPB	03 Jun 2024 16:26:56	U	SMPL
PB161243BS - 1	LCSW	17644	4.3246	-	7470A	PPB	03 Jun 2024 16:29:14	U	SMPL
P2639-01 - 1	PCB05282024RB01	495	0.1697	-	7470A	PPB	03 Jun 2024 16:31:31	U	SMPL
P2646-04 - 1	WCS-TPS	745	0.2303	-	7470A	PPB	03 Jun 2024 16:33:49	U	SMPL
P2657-04 - 1	1-AB	1062	0.3071	-	7470A	PPB	03 Jun 2024 16:36:06	U	SMPL
P2657-08 - 1	2-AB	993	0.2904	-	7470A	PPB	03 Jun 2024 16:38:23	U	SMPL
P2657-12 - 1	3-AB	795	0.2424	-	7470A	PPB	03 Jun 2024 16:40:38	U	SMPL
CCV74 - 1	CCV74	21714	5.3106	-	7470A	PPB	03 Jun 2024 16:42:53	U	SMPL
CCB74 - 1	CCB74	-261	-0.0134	-	7470A	PPB	03 Jun 2024 16:45:09	U	SMPL
P2657-16 - 1	4-AB	956	0.2814	-	7470A	PPB	03 Jun 2024 16:47:27	U	SMPL
P2657-20 - 1	5-AB	845	0.2545	-	7470A	PPB	03 Jun 2024 16:49:43	U	SMPL
P2662-04 - 1	WCS-TPR	677	0.2138	-	7470A	PPB	03 Jun 2024 16:51:57	U	SMPL
P2662-04DUP - 1	WCS-TPRDUP	667	0.2114	-	7470A	PPB	03 Jun 2024 16:54:13	U	SMPL
P2662-04MS - 1	WCS-TPRMS	17430	4.2727	-	7470A	PPB	03 Jun 2024 16:56:29	U	SMPL
P2662-04MSD - 1	WCS-TPRMSD	17681	4.3335	-	7470A	PPB	03 Jun 2024 16:58:45	U	SMPL
PB161189TB - 1	PB161189TB	-250	-0.0108	-	7470A	PPB	03 Jun 2024 17:01:03	U	SMPL
P2662-04LX5 - 1		3	0.0505	-	7470A	PPB	03 Jun 2024 17:05:58	U	SMPL
P2662-04A - 1		16817	4.1242	-	7470A	PPB	03 Jun 2024 17:08:15	U	SMPL
CCV75 - 1	CCV75	21911	5.3584	-	7470A	PPB	03 Jun 2024 17:10:31	U	SMPL
CCB75 - 1	CCB75	-311	-0.0256	-	7470A	PPB	03 Jun 2024 17:12:48	U	SMPL